

ABX Minotrol CRP



LOT

MC 403

CONTROL



(Exp.) 2017-07-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio															
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	
				EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP			
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	1.9		2.0		± 0.4	7.9		8.3		± 1.0	20.6		22.2		± 2.2		
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.28		2.28		± 0.15	4.67		4.70		± 0.18	5.62		5.71		± 0.25		
HB	HGB	g/dl	5.9		5.7		± 0.4	13.4		13.5		± 0.5	17.6		18.0		± 0.7		
		g/l	59		57		± 4	134		135		± 5	176		180		± 7		
		mmol/l	3.66		3.54		± 0.25	8.32		8.38		± 0.31	10.93		11.18		± 0.44		
HT	HCT	%	16.0		16.0		± 2.0	37.8		37.1		± 2.5	49.5		49.1		± 3.0		
		l/l	0.160		0.160		± 0.020	0.378		0.371		± 0.025	0.495		0.491		± 0.030		
VGM	MCV	µm ³ ; fl	70		70.0		± 4	81		79.0		± 4	88		86.0		± 4		
TGMH	MCH	pg	25.9		25.0		± 2.4	28.7		28.7		± 2.4	31.3		31.5		± 2.8		
		fmol	1.61		1.55		± 0.15	1.78		1.78		± 0.15	1.94		1.96		± 0.18		
CCMH	MCHC	g/dl	37.0		35.7		± 3.0	35.4		36.4		± 3.0	35.6		36.7		± 3.0		
		g/l	370		357		± 30	354		364		± 30	356		367		± 30		
		mmol/l	22.96		22.18		± 1.86	22.00		22.58		± 1.86	22.10		22.76		± 1.86		
IDR	RDW	%	13.0		14.0		± 3.0	13.2		13.5		± 3.0	12.5		12.7		± 3.0		
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	75		68		± 20	268		270		± 40	520		530		± 60		
VPM	MPV	µm ³ ; fl	8.4		9.4		± 2.0	8.4		9.0		± 2.0	8.4		8.9		± 2.0		
LYMPHO		#	1.2		N/A		± 0.3	2.8		N/A		± 0.6	3.7		N/A		± 1.3		
		%	62.0		N/A		± 7.0	36.0		N/A		± 6.0	18.0		N/A		± 6.0		
MONO		#	0.2		N/A		± 0.2	0.6		N/A		± 0.4	1.0		N/A		± 1.0		
		%	8.0		N/A		± 6.0	8.0		N/A		± 5.0	5.0		N/A		± 5.0		
GRANULO		#	0.6		N/A		± 0.3	4.4		N/A		± 0.6	15.9		N/A		± 1.5		
		%	30.0		N/A		± 8.0	56.0		N/A		± 6.0	77.0		N/A		± 7.0		
CRP		mg/dl	0.81		0.78		± 0.40	2.52		2.62		± 0.80	3.87		4.18		± 1.50		
		mg/l	8.1		7.8		± 4.0	25.2		26.2		± 8.0	38.7		41.8		± 15.0		

Ref: TEMP-1070 Rev.37 BACKVERSO 9930089-A