

ABX Minotrol CRP



LOT

MC 413

CONTROL



(Exp.) 2020-01-05
(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio															
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA			
				EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP			
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1		2.1		± 0.4	7.8		8.1		± 1.0	20.4		21.4		± 2.2		
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.39		2.33		± 0.15	4.54		4.55		± 0.18	5.55		5.61		± 0.25		
HB	HGB	g/dl	6.2		5.9		± 0.4	13.3		13.4		± 0.5	17.4		17.7		± 0.7		
		g/l	62		59		± 4	133		134		± 5	174		177		± 7		
		mmol/l	3.85		3.66		± 0.25	8.26		8.32		± 0.31	10.81		10.99		± 0.44		
HT	HCT	%	17.0		16.4		± 2.0	36.3		36.4		± 2.5	48.3		47.7		± 3.0		
		l/l	0.170		0.164		± 0.020	0.363		0.364		± 0.025	0.483		0.477		± 0.030		
VGM	MCV	µm ³ ; fl	71		70.5		± 4	80		80.0		± 4	87		85.0		± 4		
TGMH	MCH	pg	25.9		25.3		± 2.4	29.3		29.5		± 2.4	31.4		31.6		± 2.8		
		fmol	1.61		1.57		± 0.15	1.82		1.83		± 0.15	1.95		1.96		± 0.18		
CCMH	MCHC	g/dl	36.5		35.9		± 3.0	36.6		36.8		± 3.0	36.0		37.1		± 3.0		
		g/l	365		359		± 30	366		368		± 30	360		371		± 30		
		mmol/l	22.69		22.30		± 1.86	22.74		22.86		± 1.86	22.38		23.05		± 1.86		
IDR	RDW	%	13.5		14.5		± 3.0	12.8		13.0		± 3.0	12.0		12.0		± 3.0		
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	70		68		± 20	258		260		± 40	480		513		± 60		
VPM	MPV	µm ³ ; fl	7.6		9.0		± 2.0	8.2		8.8		± 2.0	7.9		8.2		± 2.0		
LYMPHO		#	1.4		N/A		± 0.2	2.7		N/A		± 0.5	3.9		N/A		± 1.3		
		%	66.5		N/A		± 7.0	34.5		N/A		± 6.0	19.0		N/A		± 6.0		
MONO		#	0.2		N/A		± 0.2	0.7		N/A		± 0.4	1.1		N/A		± 0.8		
		%	7.5		N/A		± 6.0	8.5		N/A		± 5.0	5.5		N/A		± 4.0		
GRANULO		#	0.5		N/A		± 0.2	4.4		N/A		± 0.5	15.4		N/A		± 1.5		
		%	26.0		N/A		± 7.0	57.0		N/A		± 6.0	75.5		N/A		± 7.0		
CRP		mg/dl	0.49		0.44		± 0.40	2.33		2.01		± 0.80	4.28		3.89		± 1.50		
		mg/l	4.9		4.4		± 4.0	23.3		20.1		± 8.0	42.8		38.9		± 15.0		

Ref: TEMP-1070 Rev.38 BACK/VERSO 1300060514