

ABX Minotrol CRP



LOT

MC 423

CONTROL



(Exp.) 2022-07-05
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE	
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA				
				EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G			MS CRP		EMI CRP	LC-767G			MS CRP
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.2	2.1		2.2		± 0.4	7.6	7.6		8.0		± 1.0	20.7	21.1		21.6		± 2.2
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.24	2.23		2.22		± 0.15	4.50	4.50		4.54		± 0.18	5.56	5.62		5.66		± 0.25
HB	HGB	g/dl	6.0	6.0		5.8		± 0.4	13.2	13.3		13.4		± 0.5	18.0	18.0		18.4		± 0.7
		g/l	60	60		58		± 4	132	133		134		± 5	180	180		184		± 7
		mmol/l	3.73	3.73		3.60		± 0.25	8.20	8.26		8.32		± 0.31	11.18	11.18		11.43		± 0.44
HT	HCT	%	16.1	16.5		16.0		± 2.0	36.5	36.5		36.3		± 2.5	49.5	50.0		49.8		± 3.0
		l/l	0.161	0.165		0.160		± 0.020	0.365	0.365		0.363		± 0.025	0.495	0.500		0.498		± 0.030
VGM	MCV	µm ³ ; fl	72	74		72.0		± 4	81	81		80.0		± 4	89	89		88.0		± 4
TGMH	MCH	pg	26.8	26.9		26.1		± 2.4	29.3	29.6		29.5		± 2.4	32.4	32.0		32.5		± 2.8
		fmol	1.66	1.67		1.62		± 0.15	1.82	1.84		1.83		± 0.15	2.01	1.99		2.02		± 0.18
CCMH	MCHC	g/dl	37.2	36.4		36.3		± 3.0	36.2	36.5		36.9		± 3.0	36.4	36.0		36.9		± 3.0
		g/l	372	364		363		± 30	362	365		369		± 30	364	360		369		± 30
		mmol/l	23.10	22.58		22.53		± 1.86	22.49	22.66		22.91		± 1.86	22.59	22.35		22.94		± 1.86
IDR	RDW	%	13.0	13.0		15.0		± 3.0	12.5	12.5		12.5		± 3.0	12.0	12.0		12.0		± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	75	82		68		± 20	254	260		252		± 40	489	498		495		± 60
VPM	MPV	µm ³ ; fl	8.0	8.7		9.4		± 2.0	8.5	8.6		9.0		± 2.0	8.5	8.6		9.0		± 2.0
LYMPHO		#	1.3	1.2		N/A		± 0.2	2.5	2.4		N/A		± 0.5	3.7	3.6		N/A		± 1.1
		%	59.0	59.0		N/A		± 8.0	33.0	32.0		N/A		± 6.0	18.0	17.0		N/A		± 5.0
MONO		#	0.2	0.2		N/A		± 0.2	0.6	0.6		N/A		± 0.4	1.2	1.3		N/A		± 1.0
		%	10.5	11.0		N/A		± 7.0	8.0	8.0		N/A		± 5.0	6.0	6.0		N/A		± 5.0
GRANULO		#	0.7	0.6		N/A		± 0.2	4.5	4.6		N/A		± 0.5	15.7	16.2		N/A		± 1.5
		%	30.5	30.0		N/A		± 7.0	59.0	60.0		N/A		± 6.0	76.0	77.0		N/A		± 7.0
CRP		mg/dl	0.46	0.52		0.57		± 0.30	2.66	2.77		2.69		± 0.70	4.00	4.04		4.23		± 1.00
		mg/l	4.6	5.2		5.7		± 3.0	26.6	27.7		26.9		± 7.0	40.0	40.4		42.3		± 10.0

Ref: TEMP-1070 Rev.39 BACK/VERSO 1300092170