

# ABX Minotrol CRP



**LOT**

**MC 414**

**CONTROL**



**(Exp.) 2020-04-05**  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL					1	CONTROL					2	CONTROL					3
		MICROS			PENTRA		TOLERANCES TOLERANCE	MICROS			PENTRA		TOLERANCES TOLERANCE	MICROS			PENTRA		TOLERANCES TOLERANCE
		EMI CRP			MS CRP			EMI CRP			MS CRP			EMI CRP			MS CRP		
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>1.9</b>			<b>1.9</b>		± 0.4	<b>7.9</b>			<b>8.1</b>		± 1.0	<b>20.6</b>			<b>21.5</b>		± 2.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.48</b>			<b>2.41</b>		± 0.15	<b>4.60</b>			<b>4.62</b>		± 0.18	<b>5.60</b>			<b>5.67</b>		± 0.25
HB HGB	g/dl	<b>6.2</b>			<b>6.0</b>		± 0.4	<b>13.4</b>			<b>13.6</b>		± 0.5	<b>17.6</b>			<b>18.1</b>		± 0.7
	g/l	<b>62</b>			<b>60</b>		± 4	<b>134</b>			<b>136</b>		± 5	<b>176</b>			<b>181</b>		± 7
	mmol/l	<b>3.85</b>			<b>3.73</b>		± 0.25	<b>8.32</b>			<b>8.45</b>		± 0.31	<b>10.93</b>			<b>11.24</b>		± 0.44
HT HCT	%	<b>17.1</b>			<b>16.9</b>		± 2.0	<b>37.3</b>			<b>37.0</b>		± 2.5	<b>49.3</b>			<b>48.8</b>		± 3.0
	l/l	<b>0.171</b>			<b>0.169</b>		± 0.020	<b>0.373</b>			<b>0.370</b>		± 0.025	<b>0.493</b>			<b>0.488</b>		± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>69</b>			<b>70.0</b>		± 4	<b>81</b>			<b>80.0</b>		± 4	<b>88</b>			<b>86.0</b>		± 4
TGMH MCH	pg	<b>25.0</b>			<b>24.9</b>		± 2.4	<b>29.1</b>			<b>29.4</b>		± 2.4	<b>31.4</b>			<b>31.9</b>		± 2.8
	fmol	<b>1.55</b>			<b>1.55</b>		± 0.15	<b>1.81</b>			<b>1.83</b>		± 0.15	<b>1.95</b>			<b>1.98</b>		± 0.18
CCMH MCHC	g/dl	<b>36.2</b>			<b>35.6</b>		± 3.0	<b>36.0</b>			<b>36.8</b>		± 3.0	<b>35.7</b>			<b>37.1</b>		± 3.0
	g/l	<b>362</b>			<b>356</b>		± 30	<b>360</b>			<b>368</b>		± 30	<b>357</b>			<b>371</b>		± 30
	mmol/l	<b>22.50</b>			<b>22.09</b>		± 1.86	<b>22.33</b>			<b>22.85</b>		± 1.86	<b>22.18</b>			<b>23.05</b>		± 1.86
IDR RDW	%	<b>14.0</b>			<b>14.5</b>		± 3.0	<b>13.5</b>			<b>13.0</b>		± 3.0	<b>13.0</b>			<b>12.5</b>		± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>66</b>			<b>65</b>		± 20	<b>260</b>			<b>264</b>		± 40	<b>555</b>			<b>578</b>		± 60
VPM MPV	µm <sup>3</sup> ; fl	<b>7.9</b>			<b>9.0</b>		± 2.0	<b>8.3</b>			<b>8.8</b>		± 2.0	<b>8.4</b>			<b>8.5</b>		± 2.0
LYMPHO	#	<b>1.2</b>			N/A		± 0.2	<b>2.8</b>			N/A		± 0.5	<b>3.6</b>			N/A		± 1.3
	%	<b>64.5</b>			N/A		± 8.0	<b>35.0</b>			N/A		± 6.0	<b>17.5</b>			N/A		± 6.0
MONO	#	<b>0.2</b>			N/A		± 0.2	<b>0.5</b>			N/A		± 0.4	<b>1.0</b>			N/A		± 0.8
	%	<b>8.0</b>			N/A		± 7.0	<b>6.0</b>			N/A		± 5.0	<b>5.0</b>			N/A		± 4.0
GRANULO	#	<b>0.5</b>			N/A		± 0.2	<b>4.7</b>			N/A		± 0.5	<b>16.0</b>			N/A		± 1.5
	%	<b>27.5</b>			N/A		± 7.0	<b>59.0</b>			N/A		± 6.0	<b>77.5</b>			N/A		± 7.0
CRP	mg/dl	<b>0.62</b>			<b>0.57</b>		± 0.40	<b>2.95</b>			<b>2.60</b>		± 0.80	<b>4.79</b>			<b>4.48</b>		± 1.50
	mg/l	<b>6.2</b>			<b>5.7</b>		± 4.0	<b>29.5</b>			<b>26.0</b>		± 8.0	<b>47.9</b>			<b>44.8</b>		± 15.0

Ref: TEMP-1070 Rev.38 BACK/VERSO 1300060514