

ABX Minotrol CRP



LOT

MC 405

CONTROL



(Exp.) 2018-01-05

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio												TOLERANCES TOLERANCE	
				CONTROL				CONTROL				CONTROL					
				MICROS		PENTRA		MICROS		PENTRA		MICROS		PENTRA			
				EMI CRP		MS CRP		EMI CRP		MS CRP		EMI CRP		MS CRP			
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.2		2.3		± 0.4	7.9		8.2		± 1.0	20.4		21.7		± 2.2
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.26		2.24		± 0.15	4.62		4.68		± 0.18	5.55		5.65		± 0.25
HB	HGB	g/dl	5.9		5.6		± 0.4	13.5		13.6		± 0.5	17.6		18.1		± 0.7
		g/l	59		56		± 4	135		136		± 5	176		181		± 7
		mmol/l	3.66		3.48		± 0.25	8.38		8.45		± 0.31	10.93		11.24		± 0.44
HT	HCT	%	15.6		15.5		± 2.0	37.4		37.0		± 2.5	48.8		48.6		± 3.0
		l/l	0.156		0.155		± 0.020	0.374		0.370		± 0.025	0.488		0.486		± 0.030
VGM	MCV	µm ³ ; fl	69		69.0		± 4	81		79.0		± 4	88		86.0		± 4
TGMH	MCH	pg	26.1		25.0		± 2.4	29.2		29.1		± 2.4	31.7		32.0		± 2.8
		fmol	1.62		1.55		± 0.15	1.81		1.80		± 0.15	1.97		1.99		± 0.18
CCMH	MCHC	g/dl	37.8		36.2		± 3.0	36.1		36.8		± 3.0	36.0		37.3		± 3.0
		g/l	378		362		± 30	361		368		± 30	360		373		± 30
		mmol/l	23.50		22.50		± 1.86	22.40		22.84		± 1.86	22.38		23.13		± 1.86
IDR	RDW	%	13.0		13.7		± 3.0	13.0		13.0		± 3.0	12.5		12.0		± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	75		63		± 20	260		255		± 40	535		525		± 60
VPM	MPV	µm ³ ; fl	8.2		9.4		± 2.0	8.4		8.9		± 2.0	8.7		8.8		± 2.0
LYMPHO		#	1.4		N/A		± 0.2	2.8		N/A		± 0.5	3.7		N/A		± 1.3
		%	64.5		N/A		± 8.0	36.0		N/A		± 6.0	18.0		N/A		± 6.0
MONO		#	0.2		N/A		± 0.2	0.5		N/A		± 0.5	1.0		N/A		± 1.0
		%	7.5		N/A		± 7.0	6.0		N/A		± 5.0	5.0		N/A		± 5.0
GRANULO		#	0.6		N/A		± 0.2	4.6		N/A		± 0.5	15.7		N/A		± 1.5
		%	28.0		N/A		± 7.0	58.0		N/A		± 6.0	77.0		N/A		± 7.0
CRP		mg/dl	0.51		0.49		± 0.40	2.06		1.79		± 0.80	3.43		3.10		± 1.50
		mg/l	5.1		4.9		± 4.0	20.6		17.9		± 8.0	34.3		31.0		± 15.0

Ref: TEMP-1070 Rev.37 BACKVERSO 9930089-A