

# ABX Minotrol CRP



**LOT**

**MC 415**

**CONTROL**



**(Exp.) 2020-07-05**  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														
				CONTROL		1	CONTROL		2	CONTROL		3	TOLERANCES TOLERANCE					
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
				EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP		
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>		<b>2.1</b>		± 0.4	<b>7.7</b>		<b>8.2</b>		± 1.0	<b>20.9</b>		<b>21.9</b>		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.44</b>		<b>2.47</b>		± 0.15	<b>4.65</b>		<b>4.70</b>		± 0.18	<b>5.57</b>		<b>5.67</b>		± 0.25	
HB	HGB	g/dl	<b>6.3</b>		<b>6.1</b>		± 0.4	<b>13.4</b>		<b>13.6</b>		± 0.5	<b>18.0</b>		<b>18.5</b>		± 0.7	
		g/l	<b>63</b>		<b>61</b>		± 4	<b>134</b>		<b>136</b>		± 5	<b>180</b>		<b>185</b>		± 7	
		mmol/l	<b>3.91</b>		<b>3.79</b>		± 0.25	<b>8.32</b>		<b>8.45</b>		± 0.31	<b>11.18</b>		<b>11.49</b>		± 0.44	
HT	HCT	%	<b>16.8</b>		<b>17.0</b>		± 2.0	<b>37.2</b>		<b>37.1</b>		± 2.5	<b>50.1</b>		<b>49.9</b>		± 3.0	
		l/l	<b>0.168</b>		<b>0.170</b>		± 0.020	<b>0.372</b>		<b>0.371</b>		± 0.025	<b>0.501</b>		<b>0.499</b>		± 0.030	
VGM	MCV	µm <sup>3</sup> ; fl	<b>69</b>		<b>69.0</b>		± 4	<b>80</b>		<b>79.0</b>		± 4	<b>90</b>		<b>88.0</b>		± 4	
TGMH	MCH	pg	<b>25.8</b>		<b>24.7</b>		± 2.4	<b>28.8</b>		<b>28.9</b>		± 2.4	<b>32.3</b>		<b>32.6</b>		± 2.8	
		fmol	<b>1.60</b>		<b>1.53</b>		± 0.15	<b>1.79</b>		<b>1.80</b>		± 0.15	<b>2.01</b>		<b>2.03</b>		± 0.18	
CCMH	MCHC	g/dl	<b>37.4</b>		<b>35.8</b>		± 3.0	<b>36.0</b>		<b>36.6</b>		± 3.0	<b>35.9</b>		<b>37.1</b>		± 3.0	
		g/l	<b>374</b>		<b>358</b>		± 30	<b>360</b>		<b>366</b>		± 30	<b>359</b>		<b>371</b>		± 30	
		mmol/l	<b>23.24</b>		<b>22.23</b>		± 1.86	<b>22.37</b>		<b>22.75</b>		± 1.86	<b>22.30</b>		<b>23.02</b>		± 1.86	
IDR	RDW	%	<b>13.4</b>		<b>14.7</b>		± 3.0	<b>13.1</b>		<b>13.5</b>		± 3.0	<b>12.7</b>		<b>12.1</b>		± 3.0	
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>60</b>		<b>60</b>		± 20	<b>246</b>		<b>250</b>		± 40	<b>495</b>		<b>510</b>		± 60	
VPM	MPV	µm <sup>3</sup> ; fl	<b>8.0</b>		<b>9.2</b>		± 2.0	<b>8.1</b>		<b>8.7</b>		± 2.0	<b>8.2</b>		<b>8.6</b>		± 2.0	
LYMPHO		#	<b>1.3</b>		N/A		± 0.2	<b>2.7</b>		N/A		± 0.5	<b>3.7</b>		N/A		± 1.3	
		%	<b>63.0</b>		N/A		± 8.0	<b>35.0</b>		N/A		± 6.0	<b>17.5</b>		N/A		± 6.0	
MONO		#	<b>0.2</b>		N/A		± 0.2	<b>0.5</b>		N/A		± 0.4	<b>1.0</b>		N/A		± 0.8	
		%	<b>8.0</b>		N/A		± 7.0	<b>6.0</b>		N/A		± 5.0	<b>5.0</b>		N/A		± 4.0	
GRANULO		#	<b>0.6</b>		N/A		± 0.2	<b>4.5</b>		N/A		± 0.5	<b>16.2</b>		N/A		± 1.5	
		%	<b>29.0</b>		N/A		± 7.0	<b>59.0</b>		N/A		± 6.0	<b>77.5</b>		N/A		± 7.0	
CRP		mg/dl	<b>0.68</b>		<b>0.63</b>		± 0.40	<b>2.49</b>		<b>2.23</b>		± 0.80	<b>5.44</b>		<b>4.97</b>		± 1.50	
		mg/l	<b>6.8</b>		<b>6.3</b>		± 4.0	<b>24.9</b>		<b>22.3</b>		± 8.0	<b>54.4</b>		<b>49.7</b>		± 15.0	

Ref: TEMP-1070 Rev.38 BACK/VERSO 1300060514