

# ABX Minotrol CRP



**LOT**

**MC 406**

**CONTROL**



**(Exp.) 2018-04-05**  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														
				CONTROL		1	CONTROL		2	CONTROL		3	TOLERANCES TOLERANCE					
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
				EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP		
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>		<b>2.2</b>		± 0.4	<b>7.9</b>		<b>8.2</b>		± 1.0	<b>20.8</b>		<b>21.7</b>		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.34</b>		<b>2.32</b>		± 0.15	<b>4.72</b>		<b>4.72</b>		± 0.18	<b>5.65</b>		<b>5.72</b>		± 0.25	
HB	HGB	g/dl	<b>6.0</b>		<b>5.8</b>		± 0.4	<b>13.5</b>		<b>13.7</b>		± 0.5	<b>17.5</b>		<b>18.1</b>		± 0.7	
		g/l	<b>60</b>		<b>58</b>		± 4	<b>135</b>		<b>137</b>		± 5	<b>175</b>		<b>181</b>		± 7	
		mmol/l	<b>3.73</b>		<b>3.60</b>		± 0.25	<b>8.38</b>		<b>8.51</b>		± 0.31	<b>10.87</b>		<b>11.24</b>		± 0.44	
HT	HCT	%	<b>16.4</b>		<b>16.2</b>		± 2.0	<b>38.2</b>		<b>37.3</b>		± 2.5	<b>49.2</b>		<b>49.2</b>		± 3.0	
		l/l	<b>0.164</b>		<b>0.162</b>		± 0.020	<b>0.382</b>		<b>0.373</b>		± 0.025	<b>0.492</b>		<b>0.492</b>		± 0.030	
VGM	MCV	µm <sup>3</sup> ; fl	<b>70</b>		<b>70.0</b>		± 4	<b>81</b>		<b>79.0</b>		± 4	<b>87</b>		<b>86.0</b>		± 4	
TGMH	MCH	pg	<b>25.6</b>		<b>25.0</b>		± 2.4	<b>28.6</b>		<b>29.0</b>		± 2.4	<b>31.0</b>		<b>31.6</b>		± 2.8	
		fmol	<b>1.59</b>		<b>1.55</b>		± 0.15	<b>1.78</b>		<b>1.80</b>		± 0.15	<b>1.92</b>		<b>1.97</b>		± 0.18	
CCMH	MCHC	g/dl	<b>36.6</b>		<b>35.7</b>		± 3.0	<b>35.3</b>		<b>36.7</b>		± 3.0	<b>35.6</b>		<b>36.8</b>		± 3.0	
		g/l	<b>366</b>		<b>357</b>		± 30	<b>353</b>		<b>367</b>		± 30	<b>356</b>		<b>368</b>		± 30	
		mmol/l	<b>22.75</b>		<b>22.18</b>		± 1.86	<b>21.93</b>		<b>22.82</b>		± 1.86	<b>22.11</b>		<b>22.85</b>		± 1.86	
IDR	RDW	%	<b>13.5</b>		<b>14.0</b>		± 3.0	<b>13.0</b>		<b>13.5</b>		± 3.0	<b>12.0</b>		<b>13.0</b>		± 3.0	
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>75</b>		<b>73</b>		± 20	<b>260</b>		<b>250</b>		± 40	<b>495</b>		<b>510</b>		± 60	
VPM	MPV	µm <sup>3</sup> ; fl	<b>7.8</b>		<b>9.0</b>		± 2.0	<b>8.6</b>		<b>9.2</b>		± 2.0	<b>8.1</b>		<b>8.4</b>		± 2.0	
LYMPHO		#	<b>1.4</b>		N/A		± 0.2	<b>2.8</b>		N/A		± 0.5	<b>3.7</b>		N/A		± 1.3	
		%	<b>65.0</b>		N/A		± 8.0	<b>36.0</b>		N/A		± 6.0	<b>18.0</b>		N/A		± 6.0	
MONO		#	<b>0.1</b>		N/A		± 0.1	<b>0.5</b>		N/A		± 0.5	<b>0.9</b>		N/A		± 0.9	
		%	<b>6.5</b>		N/A		± 6.0	<b>6.0</b>		N/A		± 5.0	<b>4.5</b>		N/A		± 4.5	
GRANULO		#	<b>0.6</b>		N/A		± 0.2	<b>4.6</b>		N/A		± 0.5	<b>16.1</b>		N/A		± 1.5	
		%	<b>28.5</b>		N/A		± 7.0	<b>58.0</b>		N/A		± 6.0	<b>77.5</b>		N/A		± 7.0	
CRP		mg/dl	<b>0.66</b>		<b>0.62</b>		± 0.40	<b>2.38</b>		<b>2.12</b>		± 0.80	<b>4.66</b>		<b>4.27</b>		± 1.50	
		mg/l	<b>6.6</b>		<b>6.2</b>		± 4.0	<b>23.8</b>		<b>21.2</b>		± 8.0	<b>46.6</b>		<b>42.7</b>		± 15.0	

Ref: TEMP-1070 Rev.37 BACK/VERSO 9930089-A