

ABX Minotrol CRP



LOT

MC 407

CONTROL



(Exp.) 2018-07-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														
				CONTROL				1	CONTROL				2	CONTROL				3
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
				EMI CRP			MS CRP			EMI CRP				MS CRP		EMI CRP		
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0		2.1		± 0.4	7.8		8.2		± 1.0	20.5		21.8		± 2.2	
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.36		2.35		± 0.15	4.61		4.66		± 0.18	5.80		5.90		± 0.25	
HB	HGB	g/dl	6.2		6.0		± 0.4	13.3		13.6		± 0.5	17.6		18.2		± 0.7	
		g/l	62		60		± 4	133		136		± 5	176		182		± 7	
		mmol/l	3.85		3.73		± 0.25	8.26		8.45		± 0.31	10.93		11.30		± 0.44	
HT	HCT	%	16.5		16.5		± 2.0	36.9		37.0		± 2.5	49.3		49.0		± 3.0	
		l/l	0.165		0.165		± 0.020	0.369		0.370		± 0.025	0.493		0.490		± 0.030	
VGM	MCV	µm ³ ; fl	70		70.0		± 4	80		79.5		± 4	85		83.0		± 4	
TGMH	MCH	pg	26.3		25.5		± 2.4	28.9		29.2		± 2.4	30.3		30.8		± 2.8	
		fmol	1.63		1.59		± 0.15	1.79		1.81		± 0.15	1.88		1.92		± 0.18	
CCMH	MCHC	g/dl	37.5		36.5		± 3.0	36.1		36.7		± 3.0	35.7		37.2		± 3.0	
		g/l	375		365		± 30	361		367		± 30	357		372		± 30	
		mmol/l	23.31		22.65		± 1.86	22.40		22.80		± 1.86	22.17		23.08		± 1.86	
IDR	RDW	%	12.7		13.5		± 3.0	12.8		13.5		± 3.0	12.0		11.5		± 3.0	
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	85		78		± 20	265		257		± 40	510		520		± 60	
VPM	MPV	µm ³ ; fl	7.6		8.2		± 2.0	8.6		9.0		± 2.0	8.5		8.4		± 2.0	
LYMPHO	#		1.3		N/A		± 0.2	2.7		N/A		± 0.5	3.3		N/A		± 1.2	
	%		67.0		N/A		± 8.0	34.0		N/A		± 6.0	16.0		N/A		± 6.0	
MONO	#		0.2		N/A		± 0.2	0.5		N/A		± 0.4	0.9		N/A		± 0.8	
	%		8.0		N/A		± 7.0	6.0		N/A		± 5.0	4.5		N/A		± 4.0	
GRANULO	#		0.5		N/A		± 0.2	4.7		N/A		± 0.5	16.3		N/A		± 1.5	
	%		25.0		N/A		± 7.0	60.0		N/A		± 6.0	79.5		N/A		± 7.0	
CRP	mg/dl		0.57		0.53		± 0.40	3.14		2.80		± 0.80	4.45		4.01		± 1.50	
	mg/l		5.7		5.3		± 4.0	31.4		28.0		± 8.0	44.5		40.1		± 15.0	

Ref: TEMP-1070 Rev.37 BACK/VERSO 9930089-A