

**HORIBA ABX SAS**  
HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée - BP 7290  
34184 MONTPELLIER Cedex 4  
FRANCE  
Tél.: 33 (0) 4 67 14 15 16  
Fax.: 33 (0) 4 67 14 15 17

**LOT**

**MC 074**

**CONTROL**



**(Exp.) 2014-10-05**

Rev 1

(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio														
			CONTROL		1		TOLERANCES TOLERANCE	CONTROL		2		TOLERANCES TOLERANCE	CONTROL		3		TOLERANCES TOLERANCE
			MICROS		PENTRA			MICROS		PENTRA			MICROS		PENTRA		
			EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP		
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>		<b>2.0</b>		± 0.4	<b>7.8</b>		<b>7.8</b>		± 1.0	<b>20.2</b>		<b>20.8</b>		± 2.2	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.38</b>		<b>2.31</b>		± 0.15	<b>4.62</b>		<b>4.53</b>		± 0.18	<b>5.45</b>		<b>5.37</b>		± 0.25	
HB HGB	g/dl	<b>6.0</b>		<b>5.7</b>		± 0.4	<b>13.4</b>		<b>13.3</b>		± 0.5	<b>17.5</b>		<b>17.7</b>		± 0.7	
	g/l	<b>60</b>		<b>57</b>		± 4	<b>134</b>		<b>133</b>		± 5	<b>175</b>		<b>177</b>		± 7	
	mmol/l	<b>3.73</b>		<b>3.54</b>		± 0.25	<b>8.32</b>		<b>8.26</b>		± 0.31	<b>10.87</b>		<b>10.99</b>		± 0.44	
HT HCT	%	<b>16.4</b>		<b>16.5</b>		± 2.0	<b>37.0</b>		<b>36.5</b>		± 2.5	<b>48.5</b>		<b>48.1</b>		± 3.0	
	l/l	<b>0.164</b>		<b>0.165</b>		± 0.020	<b>0.370</b>		<b>0.365</b>		± 0.025	<b>0.485</b>		<b>0.481</b>		± 0.030	
VGM MCV	µm <sup>3</sup> ; fl	<b>69</b>		<b>71.4</b>		± 4	<b>80</b>		<b>80.6</b>		± 4	<b>89</b>		<b>89.6</b>		± 4	
TGMH MCH	pg	<b>25.2</b>		<b>24.7</b>		± 2.4	<b>29.0</b>		<b>29.4</b>		± 2.4	<b>32.1</b>		<b>33.0</b>		± 2.8	
	fmol	<b>1.57</b>		<b>1.53</b>		± 0.15	<b>1.80</b>		<b>1.82</b>		± 0.15	<b>1.99</b>		<b>2.05</b>		± 0.18	
CCMH MCHC	g/dl	<b>36.5</b>		<b>34.6</b>		± 3.0	<b>36.3</b>		<b>36.4</b>		± 3.0	<b>36.1</b>		<b>36.8</b>		± 3.0	
	g/l	<b>365</b>		<b>346</b>		± 30	<b>363</b>		<b>364</b>		± 30	<b>361</b>		<b>368</b>		± 30	
	mmol/l	<b>22.69</b>		<b>21.46</b>		± 1.86	<b>22.51</b>		<b>22.62</b>		± 1.86	<b>22.40</b>		<b>22.84</b>		± 1.86	
IDR RDW	%	<b>13.5</b>		<b>11.4</b>		± 3.0	<b>12.8</b>		<b>9.8</b>		± 3.0	<b>12.0</b>		<b>9.0</b>		± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>80</b>		<b>72</b>		± 20	<b>258</b>		<b>265</b>		± 40	<b>500</b>		<b>510</b>		± 60	
VPM MPV	µm <sup>3</sup> ; fl	<b>8.1</b>		<b>9.0</b>		± 2.0	<b>7.8</b>		<b>8.4</b>		± 2.0	<b>8.0</b>		<b>8.4</b>		± 2.0	
LYMPHO	#	<b>1.3</b>		N/A		± 0.2	<b>2.7</b>		N/A		± 0.5	<b>3.4</b>		N/A		± 1.1	
	%	<b>63.5</b>		N/A		± 8.0	<b>34.5</b>		N/A		± 6.0	<b>17.0</b>		N/A		± 5.0	
MONO	#	<b>0.2</b>		N/A		± 0.2	<b>0.5</b>		N/A		± 0.4	<b>1.0</b>		N/A		± 1.0	
	%	<b>8.0</b>		N/A		± 7.0	<b>7.0</b>		N/A		± 5.0	<b>5.0</b>		N/A		± 5.0	
GRANULO	#	<b>0.6</b>		N/A		± 0.2	<b>4.6</b>		N/A		± 0.5	<b>15.8</b>		N/A		± 1.5	
	%	<b>28.5</b>		N/A		± 7.0	<b>58.5</b>		N/A		± 6.0	<b>78.0</b>		N/A		± 7.0	
CRP	mg/dl	<b>0.55</b>		<b>0.50</b>		± 0.30	<b>2.50</b>		<b>2.40</b>		± 0.70	<b>4.10</b>		<b>3.90</b>		± 1.00	
	mg/l	<b>5.5</b>		<b>5.0</b>		± 3.0	<b>25.0</b>		<b>24.0</b>		± 7.0	<b>41.0</b>		<b>39.0</b>		± 10.0	

9930089-A  
BACKVERSO  
Ref: TEMP-1070 Rev.35