

# ABX Minotrol CRP



**LOT** MC 102  
Rev 1

**CONTROL**

(Exp.) 2013-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																							
				CONTROL				1				CONTROL				2				CONTROL				3			
				MICROS				TOLERANCES TOLERANCE				MICROS				TOLERANCES TOLERANCE				MICROS				TOLERANCES TOLERANCE			
				EMICRP								EMICRP								EMICRP							
GB	WBC	$10^9/\text{mm}^3$ ; $10^9/\text{l}$	<b>2.1</b>					$\pm 0.4$	<b>7.6</b>					$\pm 1.0$	<b>20.0</b>					$\pm 2.2$							
GR	RBC	$10^9/\text{mm}^3$ ; $10^{12}/\text{l}$	<b>2.43</b>					$\pm 0.15$	<b>4.64</b>					$\pm 0.18$	<b>5.63</b>					$\pm 0.25$							
HB	HGB	g/dl	<b>6.2</b>					$\pm 0.4$	<b>13.5</b>					$\pm 0.5$	<b>18.1</b>					$\pm 0.7$							
		g/l	<b>62</b>					$\pm 4$	<b>135</b>					$\pm 5$	<b>181</b>					$\pm 7$							
		mmol/l	<b>3.85</b>					$\pm 0.25$	<b>8.38</b>					$\pm 0.31$	<b>11.24</b>					$\pm 0.44$							
HT	HCT	%	<b>17.0</b>					$\pm 2.0$	<b>38.0</b>					$\pm 2.5$	<b>50.7</b>					$\pm 3.0$							
		l/l	<b>0.170</b>					$\pm 0.020$	<b>0.380</b>					$\pm 0.025$	<b>0.507</b>					$\pm 0.030$							
VGM	MCV	$\mu\text{m}^3$ ; fl	<b>70</b>					$\pm 4$	<b>82</b>					$\pm 4$	<b>90</b>					$\pm 4$							
TGMH	MCH	pg	<b>25.5</b>					$\pm 2.4$	<b>29.1</b>					$\pm 2.4$	<b>32.1</b>					$\pm 2.8$							
		fmol	<b>1.58</b>					$\pm 0.15$	<b>1.81</b>					$\pm 0.15$	<b>2.00</b>					$\pm 0.18$							
CCMH	MCHC	g/dl	<b>36.4</b>					$\pm 3.0$	<b>35.5</b>					$\pm 3.0$	<b>35.7</b>					$\pm 3.0$							
		g/l	<b>364</b>					$\pm 30$	<b>355</b>					$\pm 30$	<b>357</b>					$\pm 30$							
		mmol/l	<b>22.63</b>					$\pm 1.86$	<b>22.03</b>					$\pm 1.86$	<b>22.18</b>					$\pm 1.86$							
IDR	RDW	%	<b>13.4</b>					$\pm 3.0$	<b>12.6</b>					$\pm 3.0$	<b>11.7</b>					$\pm 3.0$							
PLAQ.	PLTS	$10^9/\text{mm}^3$ ; $10^9/\text{l}$	<b>68</b>					$\pm 20$	<b>254</b>					$\pm 40$	<b>495</b>					$\pm 60$							
VPM	MPV	$\mu\text{m}^3$ ; fl	<b>8.7</b>					$\pm 2.0$	<b>8.4</b>					$\pm 2.0$	<b>8.1</b>					$\pm 2.0$							
LYMPHO		#	<b>1.3</b>					$\pm 0.2$	<b>2.6</b>					$\pm 0.5$	<b>3.0</b>					$\pm 1.0$							
		%	<b>63.0</b>					$\pm 8.0$	<b>33.7</b>					$\pm 6.0$	<b>15.0</b>					$\pm 5.0$							
MONO		#	<b>0.2</b>					$\pm 0.2$	<b>0.5</b>					$\pm 0.4$	<b>1.1</b>					$\pm 1.1$							
		%	<b>9.4</b>					$\pm 7.0$	<b>7.2</b>					$\pm 5.0$	<b>5.4</b>					$\pm 5.0$							
GRANULO		#	<b>0.6</b>					$\pm 0.2$	<b>4.5</b>					$\pm 0.5$	<b>15.9</b>					$\pm 1.4$							
		%	<b>27.6</b>					$\pm 7.0$	<b>59.1</b>					$\pm 6.0$	<b>79.6</b>					$\pm 7.0$							
CRP		mg/dl	<b>0.69</b>					$\pm 0.30$	<b>2.82</b>					$\pm 0.70$	<b>4.77</b>					$\pm 1.00$							
		mg/l	<b>6.9</b>					$\pm 3.0$	<b>28.2</b>					$\pm 7.0$	<b>47.7</b>					$\pm 10.0$							

9930064-A  
Ref: TEMP-1070 Rev.15