

# ABX Minotrol CRP



**LOT** MC 013

**CONTROL**



(Exp.) 2013-04-05  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																							
				CONTROL				1				CONTROL				2				CONTROL				3			
				MICROS				TOLERANCES TOLERANCE				MICROS				TOLERANCES TOLERANCE				MICROS				TOLERANCES TOLERANCE			
				EMICRP								EMICRP								EMICRP							
GB	WBC	$10^3/mm^3 ; 10^9/l$	<b>1.9</b>					$\pm 0.4$	<b>7.1</b>					$\pm 1.0$	<b>19.1</b>									$\pm 2.2$			
GR	RBC	$10^9/mm^3 ; 10^{12}/l$	<b>2.41</b>					$\pm 0.15$	<b>4.65</b>					$\pm 0.18$	<b>5.52</b>									$\pm 0.25$			
HB	HGB	g/dl	<b>6.0</b>					$\pm 0.4$	<b>13.4</b>					$\pm 0.5$	<b>17.3</b>									$\pm 0.7$			
		g/l	<b>60</b>					$\pm 4$	<b>134</b>					$\pm 5$	<b>173</b>									$\pm 7$			
		mmol/l	<b>3.73</b>					$\pm 0.25$	<b>8.32</b>					$\pm 0.31$	<b>10.74</b>									$\pm 0.44$			
HT	HCT	%	<b>16.6</b>					$\pm 2.0$	<b>39.5</b>					$\pm 2.5$	<b>50.2</b>									$\pm 3.0$			
		l/l	<b>0.166</b>					$\pm 0.020$	<b>0.395</b>					$\pm 0.025$	<b>0.502</b>									$\pm 0.030$			
VGM	MCV	$\mu m^3 ; fl$	<b>69</b>					$\pm 4$	<b>85</b>					$\pm 4$	<b>91</b>									$\pm 4$			
TGMH	MCH	pg	<b>24.9</b>					$\pm 2.4$	<b>28.8</b>					$\pm 2.4$	<b>31.3</b>									$\pm 2.8$			
		fmol	<b>1.55</b>					$\pm 0.15$	<b>1.79</b>					$\pm 0.15$	<b>1.95</b>									$\pm 0.18$			
CCMH	MCHC	g/dl	<b>36.1</b>					$\pm 3.0$	<b>33.9</b>					$\pm 3.0$	<b>34.4</b>									$\pm 3.0$			
		g/l	<b>361</b>					$\pm 30$	<b>339</b>					$\pm 30$	<b>344</b>									$\pm 30$			
		mmol/l	<b>22.41</b>					$\pm 1.86$	<b>21.05</b>					$\pm 1.86$	<b>21.39</b>									$\pm 1.86$			
IDR	RDW	%	<b>13.4</b>					$\pm 3.0$	<b>14.0</b>					$\pm 3.0$	<b>13.2</b>									$\pm 3.0$			
PLAQ.	PLTS	$10^3/mm^3 ; 10^9/l$	<b>67</b>					$\pm 20$	<b>240</b>					$\pm 40$	<b>472</b>									$\pm 60$			
VPM	MPV	$\mu m^3 ; fl$	<b>8.4</b>					$\pm 2.0$	<b>8.0</b>					$\pm 2.0$	<b>8.0</b>									$\pm 2.0$			
LYMPHO		#	<b>1.2</b>					$\pm 0.2$	<b>2.3</b>					$\pm 0.5$	<b>2.5</b>									$\pm 1.1$			
		%	<b>65.4</b>					$\pm 7.0$	<b>32.8</b>					$\pm 6.0$	<b>13.4</b>									$\pm 5.0$			
MONO		#	<b>0.1</b>					$\pm 0.1$	<b>0.6</b>					$\pm 0.4$	<b>1.1</b>									$\pm 1.0$			
		%	<b>7.7</b>					$\pm 7.0$	<b>8.0</b>					$\pm 5.0$	<b>5.6</b>									$\pm 5.0$			
GRANULO		#	<b>0.5</b>					$\pm 0.2$	<b>4.2</b>					$\pm 0.5$	<b>15.5</b>									$\pm 1.4$			
		%	<b>26.9</b>					$\pm 7.0$	<b>59.2</b>					$\pm 6.0$	<b>81.0</b>									$\pm 7.0$			
CRP		mg/dl	<b>0.67</b>					$\pm 0.30$	<b>2.54</b>					$\pm 0.70$	<b>4.53</b>									$\pm 1.00$			
		mg/l	<b>6.7</b>					$\pm 3.0$	<b>25.4</b>					$\pm 7.0$	<b>45.3</b>									$\pm 10.0$			

9930064-A  
Ref: TEMP-1070 Rev.15