

HORIBA ABX SAS
HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée - BP 7290
34184 MONTPELLIER Cedex 4
FRANCE
Tél.: 33 (0) 4 67 14 15 16
Fax.: 33 (0) 4 67 14 15 17

LOT

MC 103

CONTROL



(Exp.) 2014-01-05

(YYYY - MM - DD)

Rev 1

		ABX Lysebio														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				TOLERANCES TOLERANCE	CONTROL				TOLERANCES TOLERANCE	CONTROL				TOLERANCES TOLERANCE
		MICROS					MICROS					MICROS				
		EMI CRP					EMI CRP					EMI CRP				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0				± 0.4	7.7				± 1.0	20.1				± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.45				± 0.15	4.60				± 0.18	5.50				± 0.25
HB HGB	g/dl	6.1				± 0.4	13.4				± 0.5	17.4				± 0.7
	g/l	61				± 4	134				± 5	174				± 7
	mmol/l	3.79				± 0.25	8.32				± 0.31	10.81				± 0.44
HT HCT	%	16.9				± 2.0	37.3				± 2.5	49.0				± 3.0
	l/l	0.169				± 0.020	0.373				± 0.025	0.490				± 0.030
VGM MCV	µm ³ ; fl	69				± 4	81				± 4	89				± 4
TGMH MCH	pg	24.9				± 2.4	29.1				± 2.4	31.6				± 2.8
	fmol	1.55				± 0.15	1.81				± 0.15	1.96				± 0.18
CCMH MCHC	g/dl	36.1				± 3.0	36.0				± 3.0	35.5				± 3.0
	g/l	361				± 30	360				± 30	355				± 30
	mmol/l	22.41				± 1.86	22.33				± 1.86	22.07				± 1.86
IDR RDW	%	13.5				± 3.0	13.0				± 3.0	12.5				± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	76				± 20	240				± 40	485				± 60
VPM MPV	µm ³ ; fl	8.4				± 2.0	7.8				± 2.0	7.9				± 2.0
LYMPHO	#	1.3				± 0.2	2.7				± 0.5	3.6				± 1.2
	%	63.0				± 8.0	34.5				± 6.0	18.0				± 6.0
MONO	#	0.2				± 0.2	0.6				± 0.4	1.1				± 1.0
	%	9.0				± 7.0	8.0				± 5.0	5.5				± 5.0
GRANULO	#	0.6				± 0.2	4.4				± 0.5	15.4				± 1.5
	%	28.0				± 7.0	57.5				± 6.0	76.5				± 7.0
CRP	mg/dl	0.59				± 0.30	2.00				± 0.70	3.87				± 1.00
	mg/l	5.9				± 3.0	20.0				± 7.0	38.7				± 10.0

9930074-A
BACKVERSO
Ref: TEMP-1070 Rev.31