

ABX Minotrol CRP



LOT

MC 045

CONTROL



(Exp.) 2015-07-05
(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio												TOLERANCES TOLERANCE	
				CONTROL		1	CONTROL		2	CONTROL		3					
				MICROS		PENTRA	MICROS		PENTRA	MICROS		PENTRA					
				EMI CRP		MS CRP	EMI CRP		MS CRP	EMI CRP		MS CRP					
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0		2.0		± 0.4	7.9		7.9		± 1.0	20.6		21.0		± 2.2
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.57		2.52		± 0.15	4.61		4.60		± 0.18	5.75		5.73		± 0.25
HB	HGB	g/dl	6.2		6.1		± 0.4	13.3		13.5		± 0.5	17.8		18.3		± 0.7
		g/l	62		61		± 4	133		135		± 5	178		183		± 7
		mmol/l	3.85		3.79		± 0.25	8.26		8.38		± 0.31	11.05		11.36		± 0.44
HT	HCT	%	17.5		17.4		± 2.0	36.9		36.8		± 2.5	50.0		49.9		± 3.0
		l/l	0.175		0.174		± 0.020	0.369		0.368		± 0.025	0.500		0.499		± 0.030
VGM	MCV	µm ³ ; fl	68		69.0		± 4	80		80.0		± 4	87		87.0		± 4
TGMH	MCH	pg	24.1		24.2		± 2.4	28.9		29.3		± 2.4	31.0		31.9		± 2.8
		fmol	1.50		1.50		± 0.15	1.79		1.82		± 0.15	1.92		1.98		± 0.18
CCMH	MCHC	g/dl	35.5		35.1		± 3.0	36.1		36.7		± 3.0	35.6		36.7		± 3.0
		g/l	355		351		± 30	361		367		± 30	356		367		± 30
		mmol/l	22.03		21.79		± 1.86	22.40		22.78		± 1.86	22.10		22.80		± 1.86
IDR	RDW	%	13.5		12.0		± 3.0	13.0		11.5		± 3.0	12.5		11.0		± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	74		67		± 20	258		252		± 40	510		505		± 60
VPM	MPV	µm ³ ; fl	8.5		9.3		± 2.0	8.1		8.6		± 2.0	8.1		8.4		± 2.0
LYMPHO		#	1.3		N/A		± 0.3	2.7		N/A		± 0.5	3.7		N/A		± 1.3
		%	64.5		N/A		± 8.0	34.0		N/A		± 6.0	18.0		N/A		± 6.0
MONO		#	0.2		N/A		± 0.2	0.6		N/A		± 0.5	1.1		N/A		± 1.0
		%	9.0		N/A		± 7.0	7.5		N/A		± 5.0	5.5		N/A		± 5.0
GRANULO		#	0.5		N/A		± 0.2	4.6		N/A		± 0.6	15.8		N/A		± 1.5
		%	26.5		N/A		± 7.0	58.5		N/A		± 6.0	76.5		N/A		± 7.0
CRP		mg/dl	0.54		0.50		± 0.40	2.50		2.50		± 0.80	4.10		3.90		± 1.50
		mg/l	5.4		5.0		± 4.0	25.0		25.0		± 8.0	41.0		39.0		± 15.0

Ref: TEMP-1070 Rev.36 BACK/VERSO 9930089-A