

# ABX Minotrol CRP



**LOT**

**MC 075**

Rev 2

**CONTROL**



(Exp.) **2015-10-05**

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														
		CONTROL				1	CONTROL				2	CONTROL				3
		MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
		EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP		
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2,1</b>		<b>2,0</b>		± 0,4	<b>7,6</b>		<b>7,8</b>		± 1,0	<b>19,8</b>		<b>20,6</b>		± 2,2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2,45</b>		<b>2,38</b>		± 0,15	<b>4,60</b>		<b>4,55</b>		± 0,18	<b>5,65</b>		<b>5,60</b>		± 0,25
HB HGB	g/dl	<b>6,1</b>		<b>5,9</b>		± 0,4	<b>13,5</b>		<b>13,5</b>		± 0,5	<b>17,8</b>		<b>18,1</b>		± 0,7
	g/l	<b>61</b>		<b>59</b>		± 4	<b>135</b>		<b>135</b>		± 5	<b>178</b>		<b>181</b>		± 7
HT HCT	mmol/l	<b>3,79</b>		<b>3,66</b>		± 0,25	<b>8,38</b>		<b>8,38</b>		± 0,31	<b>11,05</b>		<b>11,24</b>		± 0,44
	%	<b>16,9</b>		<b>16,4</b>		± 2,0	<b>37,3</b>		<b>36,4</b>		± 2,5	<b>49,2</b>		<b>48,2</b>		± 3,0
VGM MCV	l/l	<b>0,169</b>		<b>0,164</b>		± 0,020	<b>0,373</b>		<b>0,364</b>		± 0,025	<b>0,492</b>		<b>0,482</b>		± 0,030
	µm <sup>3</sup> ; fl	<b>69</b>		<b>69,0</b>		± 4	<b>81</b>		<b>80,0</b>		± 4	<b>87</b>		<b>86,0</b>		± 4
TGMH MCH	pg	<b>24,9</b>		<b>24,8</b>		± 2,4	<b>29,3</b>		<b>29,7</b>		± 2,4	<b>31,5</b>		<b>32,3</b>		± 2,8
	fmol	<b>1,55</b>		<b>1,54</b>		± 0,15	<b>1,82</b>		<b>1,84</b>		± 0,15	<b>1,96</b>		<b>2,01</b>		± 0,18
CCMH MCHC	g/dl	<b>36,1</b>		<b>35,9</b>		± 3,0	<b>36,2</b>		<b>37,1</b>		± 3,0	<b>36,2</b>		<b>37,6</b>		± 3,0
	g/l	<b>361</b>		<b>359</b>		± 30	<b>362</b>		<b>371</b>		± 30	<b>362</b>		<b>376</b>		± 30
IDR RDW	mmol/l	<b>22,41</b>		<b>22,31</b>		± 1,86	<b>22,50</b>		<b>23,03</b>		± 1,86	<b>22,49</b>		<b>23,34</b>		± 1,86
	%	<b>13,5</b>		<b>12,5</b>		± 3,0	<b>12,7</b>		<b>11,5</b>		± 3,0	<b>12,4</b>		<b>11,0</b>		± 3,0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>78</b>		<b>60</b>		± 20	<b>258</b>		<b>235</b>		± 40	<b>515</b>		<b>500</b>		± 60
VPM MPV	µm <sup>3</sup> ; fl	<b>8,3</b>		<b>9,3</b>		± 2,0	<b>8,2</b>		<b>9,0</b>		± 2,0	<b>8,2</b>		<b>8,5</b>		± 2,0
LYMPHO	#	<b>1,3</b>		N/A		± 0,3	<b>2,6</b>		N/A		± 0,5	<b>3,3</b>		N/A		± 1,3
	%	<b>63,0</b>		N/A		± 8,0	<b>34,0</b>		N/A		± 6,0	<b>16,5</b>		N/A		± 6,0
MONO	#	<b>0,2</b>		N/A		± 0,2	<b>0,6</b>		N/A		± 0,5	<b>1,0</b>		N/A		± 1,0
	%	<b>11,0</b>		N/A		± 7,0	<b>8,0</b>		N/A		± 5,0	<b>5,0</b>		N/A		± 5,0
GRANULO	#	<b>0,5</b>		N/A		± 0,2	<b>4,4</b>		N/A		± 0,6	<b>15,5</b>		N/A		± 1,5
	%	<b>26,0</b>		N/A		± 7,0	<b>58,0</b>		N/A		± 6,0	<b>78,5</b>		N/A		± 7,0
CRP	mg/dl	<b>0,70</b>		<b>0,60</b>		± 0,40	<b>2,40</b>		<b>2,40</b>		± 0,80	<b>4,00</b>		<b>4,20</b>		± 1,50
	mg/l	<b>7,0</b>		<b>6,0</b>		± 4,0	<b>24,0</b>		<b>24,0</b>		± 8,0	<b>40,0</b>		<b>42,0</b>		± 15,0

Ref: TEMP-1070 Rev.36 BACKVERSO 9930089-A