

ABX Minotrol CRP



LOT

MC 105

Rev 2

CONTROL



(Exp.) **2016-01-05**

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														
		CONTROL				1	CONTROL				2	CONTROL				3
		MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
		EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP		
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1		2.2		± 0.4	8.1		8.1		± 1.0	20.3		20.9		± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.44		2.40		± 0.15	4.52		4.50		± 0.18	5.56		5.55		± 0.25
HB HGB	g/dl	6.1		6.0		± 0.4	13.4		13.4		± 0.5	17.6		17.9		± 0.7
	g/l	61		60		± 4	134		134		± 5	176		179		± 7
HT HCT	mmol/l	3.79		3.73		± 0.25	8.32		8.32		± 0.31	10.93		11.12		± 0.44
	%	17.1		16.7		± 2.0	37.1		36.0		± 2.5	49.5		48.3		± 3.0
VGM MCV	l/l	0.171		0.167		± 0.020	0.371		0.360		± 0.025	0.495		0.483		± 0.030
	µm ³ ; fl	70		69.5		± 4	82		80.0		± 4	89		87.0		± 4
TGMH MCH	pg	25.0		25.0		± 2.4	29.6		29.8		± 2.4	31.7		32.3		± 2.8
	fmol	1.55		1.55		± 0.15	1.84		1.85		± 0.15	1.97		2.00		± 0.18
CCMH MCHC	g/dl	35.7		36.0		± 3.0	36.2		37.2		± 3.0	35.6		37.1		± 3.0
	g/l	357		360		± 30	362		372		± 30	356		371		± 30
IDR RDW	mmol/l	22.18		22.34		± 1.86	22.45		23.12		± 1.86	22.09		23.02		± 1.86
	%	13.1		12.2		± 3.0	12.6		11.2		± 3.0	12.1		11.5		± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	72		58		± 20	260		245		± 40	485		480		± 60
VPM MPV	µm ³ ; fl	8.7		9.8		± 2.0	8.3		9.3		± 2.0	8.2		8.8		± 2.0
LYMPHO	#	1.3		N/A		± 0.3	2.9		N/A		± 0.5	3.7		N/A		± 1.3
	%	63.0		N/A		± 8.0	35.5		N/A		± 6.0	18.0		N/A		± 6.0
MONO	#	0.2		N/A		± 0.2	0.6		N/A		± 0.5	1.0		N/A		± 1.0
	%	10.0		N/A		± 7.0	8.0		N/A		± 5.0	5.0		N/A		± 5.0
GRANULO	#	0.6		N/A		± 0.2	4.6		N/A		± 0.6	15.6		N/A		± 1.5
	%	27.0		N/A		± 7.0	56.5		N/A		± 6.0	77.0		N/A		± 7.0
CRP	mg/dl	0.55		0.50		± 0.40	2.20		2.20		± 0.80	3.60		3.60		± 1.50
	mg/l	5.5		5.0		± 4.0	22.0		22.0		± 8.0	36.0		36.0		± 15.0

Ref: TEMP-1070 Rev.36 BACKVERSO 9930089-A