

ABX Minotrol CRP



LOT

MC 408

CONTROL



(Exp.) 2018-10-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														
				CONTROL				1	CONTROL				2	CONTROL				3
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
				EMI CRP			MS CRP			EMI CRP				MS CRP		EMI CRP		
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1		2.2		± 0.4	7.7		8.1		± 1.0	20.9		22.2		± 2.2	
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.33		2.30		± 0.15	4.55		4.63		± 0.18	5.51		5.65		± 0.25	
HB	HGB	g/dl	6.1		5.9		± 0.4	13.3		13.6		± 0.5	17.7		18.3		± 0.7	
		g/l	61		59		± 4	133		136		± 5	177		183		± 7	
		mmol/l	3.79		3.66		± 0.25	8.26		8.45		± 0.31	10.99		11.36		± 0.44	
HT	HCT	%	16.5		16.1		± 2.0	36.4		36.6		± 2.5	49.0		48.6		± 3.0	
		l/l	0.165		0.161		± 0.020	0.364		0.366		± 0.025	0.490		0.486		± 0.030	
VGM	MCV	µm ³ ; fl	71		70.0		± 4	80		79.0		± 4	89		86.0		± 4	
TGMH	MCH	pg	26.2		25.7		± 2.4	29.2		29.4		± 2.4	32.1		32.4		± 2.8	
		fmol	1.63		1.59		± 0.15	1.82		1.82		± 0.15	1.99		2.01		± 0.18	
CCMH	MCHC	g/dl	36.9		36.6		± 3.0	36.5		37.2		± 3.0	36.1		37.7		± 3.0	
		g/l	369		366		± 30	365		372		± 30	361		377		± 30	
		mmol/l	22.90		22.76		± 1.86	22.69		23.09		± 1.86	22.41		23.39		± 1.86	
IDR	RDW	%	12.6		14.0		± 3.0	13.0		13.0		± 3.0	11.8		11.8		± 3.0	
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	82		71		± 20	255		265		± 40	505		518		± 60	
VPM	MPV	µm ³ ; fl	7.4		9.1		± 2.0	7.7		8.6		± 2.0	8.0		8.6		± 2.0	
LYMPHO		#	1.4		N/A		± 0.2	2.7		N/A		± 0.5	3.4		N/A		± 1.3	
		%	65.0		N/A		± 8.0	35.5		N/A		± 6.0	16.5		N/A		± 6.0	
MONO		#	0.2		N/A		± 0.2	0.5		N/A		± 0.4	1.1		N/A		± 1.0	
		%	6.5		N/A		± 6.5	6.5		N/A		± 5.0	5.5		N/A		± 5.0	
GRANULO		#	0.6		N/A		± 0.2	4.5		N/A		± 0.5	16.3		N/A		± 1.5	
		%	28.5		N/A		± 7.0	58.0		N/A		± 6.0	78.0		N/A		± 7.0	
CRP		mg/dl	0.59		0.56		± 0.40	2.60		2.37		± 0.80	4.76		4.62		± 1.50	
		mg/l	5.9		5.6		± 4.0	26.0		23.7		± 8.0	47.6		46.2		± 15.0	

Ref: TEMP-1070 Rev.37 BACK/VERSO 9930089-A