

ABX Minotrol CRP



LOT

MC 409

CONTROL



(Exp.) 2019-01-05

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														
				CONTROL				1	CONTROL				2	CONTROL				3
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
				EMI CRP			MS CRP			EMI CRP				MS CRP		EMI CRP		
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1		2.2		± 0.4	7.7		8.2		± 1.0	20.2		21.8		± 2.2	
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.35		2.34		± 0.15	4.60		4.66		± 0.18	5.72		5.92		± 0.25	
HB	HGB	g/dl	6.2		6.1		± 0.4	13.4		13.6		± 0.5	17.8		18.6		± 0.7	
		g/l	62		61		± 4	134		136		± 5	178		186		± 7	
		mmol/l	3.85		3.79		± 0.25	8.32		8.45		± 0.31	11.05		11.55		± 0.44	
HT	HCT	%	16.7		16.6		± 2.0	36.8		37.0		± 2.5	49.8		50.0		± 3.0	
		l/l	0.167		0.166		± 0.020	0.368		0.370		± 0.025	0.498		0.500		± 0.030	
VGM	MCV	µm ³ ; fl	71		71.0		± 4	80		79.5		± 4	87		84.5		± 4	
TGMH	MCH	pg	26.4		26.1		± 2.4	29.1		29.2		± 2.4	31.1		31.4		± 2.8	
		fmol	1.64		1.62		± 0.15	1.81		1.81		± 0.15	1.93		1.95		± 0.18	
CCMH	MCHC	g/dl	37.2		36.7		± 3.0	36.4		36.7		± 3.0	35.8		37.2		± 3.0	
		g/l	372		367		± 30	364		367		± 30	358		372		± 30	
		mmol/l	23.08		22.80		± 1.86	22.61		22.80		± 1.86	22.21		23.09		± 1.86	
IDR	RDW	%	13.0		14.5		± 3.0	12.5		13.5		± 3.0	12.0		12.0		± 3.0	
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	80		73		± 20	260		255		± 40	495		510		± 60	
VPM	MPV	µm ³ ; fl	7.8		9.1		± 2.0	8.2		9.0		± 2.0	8.2		8.7		± 2.0	
LYMPHO		#	1.3		N/A		± 0.2	2.7		N/A		± 0.5	3.6		N/A		± 1.2	
		%	64.0		N/A		± 8.0	35.5		N/A		± 6.0	18.0		N/A		± 6.0	
MONO		#	0.2		N/A		± 0.2	0.5		N/A		± 0.4	1.1		N/A		± 1.0	
		%	7.0		N/A		± 7.0	7.0		N/A		± 5.0	5.5		N/A		± 5.0	
GRANULO		#	0.6		N/A		± 0.2	4.4		N/A		± 0.6	15.5		N/A		± 1.5	
		%	29.0		N/A		± 7.0	57.5		N/A		± 6.0	76.5		N/A		± 7.0	
CRP		mg/dl	0.56		0.51		± 0.30	2.50		2.11		± 0.80	4.50		3.88		± 1.50	
		mg/l	5.6		5.1		± 3.0	25.0		21.1		± 8.0	45.0		38.8		± 15.0	

Ref: TEMP-1070 Rev.37 BACK/VERSO 9930089-A