

# ABX Minotrol CRP



**LOT**

**MC 422**

**CONTROL**



**(Exp.) 2022-04-05**  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE	
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA				
				EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G			MS CRP		EMI CRP	LC-767G			MS CRP
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>	<b>2.1</b>		<b>2.2</b>		± 0.4	<b>8.7</b>	<b>8.7</b>		<b>9.1</b>		± 1.0	<b>21.5</b>	<b>21.8</b>		<b>22.5</b>		± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.36</b>	<b>2.33</b>		<b>2.32</b>		± 0.15	<b>4.50</b>	<b>4.47</b>		<b>4.50</b>		± 0.18	<b>5.46</b>	<b>5.40</b>		<b>5.49</b>		± 0.25
HB	HGB	g/dl	<b>6.2</b>	<b>6.2</b>		<b>5.9</b>		± 0.4	<b>13.3</b>	<b>13.4</b>		<b>13.4</b>		± 0.5	<b>17.3</b>	<b>17.4</b>		<b>17.8</b>		± 0.7
		g/l	<b>62</b>	<b>62</b>		<b>59</b>		± 4	<b>133</b>	<b>134</b>		<b>134</b>		± 5	<b>173</b>	<b>174</b>		<b>178</b>		± 7
		mmol/l	<b>3.85</b>	<b>3.85</b>		<b>3.66</b>		± 0.25	<b>8.26</b>	<b>8.32</b>		<b>8.32</b>		± 0.31	<b>10.74</b>	<b>10.81</b>		<b>11.05</b>		± 0.44
HT	HCT	%	<b>16.8</b>	<b>17.0</b>		<b>17.2</b>		± 2.0	<b>36.5</b>	<b>36.7</b>		<b>37.2</b>		± 2.5	<b>48.0</b>	<b>48.1</b>		<b>47.2</b>		± 3.0
		l/l	<b>0.168</b>	<b>0.170</b>		<b>0.172</b>		± 0.020	<b>0.365</b>	<b>0.367</b>		<b>0.372</b>		± 0.025	<b>0.480</b>	<b>0.481</b>		<b>0.472</b>		± 0.030
VGM	MCV	µm <sup>3</sup> ; fl	<b>71</b>	<b>73</b>		<b>74.1</b>		± 4	<b>81</b>	<b>82</b>		<b>82.7</b>		± 4	<b>88</b>	<b>89</b>		<b>86.0</b>		± 4
TGMH	MCH	pg	<b>26.3</b>	<b>26.6</b>		<b>25.4</b>		± 2.4	<b>29.6</b>	<b>30.0</b>		<b>29.8</b>		± 2.4	<b>31.7</b>	<b>32.2</b>		<b>32.4</b>		± 2.8
		fmol	<b>1.63</b>	<b>1.65</b>		<b>1.58</b>		± 0.15	<b>1.84</b>	<b>1.86</b>		<b>1.85</b>		± 0.15	<b>1.97</b>	<b>2.00</b>		<b>2.01</b>		± 0.18
CCMH	MCHC	g/dl	<b>37.0</b>	<b>36.5</b>		<b>34.3</b>		± 3.0	<b>36.5</b>	<b>36.6</b>		<b>36.0</b>		± 3.0	<b>36.0</b>	<b>36.2</b>		<b>37.7</b>		± 3.0
		g/l	<b>370</b>	<b>365</b>		<b>343</b>		± 30	<b>365</b>	<b>366</b>		<b>360</b>		± 30	<b>360</b>	<b>362</b>		<b>377</b>		± 30
		mmol/l	<b>22.98</b>	<b>22.64</b>		<b>21.31</b>		± 1.86	<b>22.66</b>	<b>22.70</b>		<b>22.36</b>		± 1.86	<b>22.36</b>	<b>22.48</b>		<b>23.41</b>		± 1.86
IDR	RDW	%	<b>13.0</b>	<b>13.0</b>		<b>15.0</b>		± 3.0	<b>13.0</b>	<b>13.0</b>		<b>13.0</b>		± 3.0	<b>12.0</b>	<b>12.0</b>		<b>11.5</b>		± 3.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>71</b>	<b>78</b>		<b>67</b>		± 20	<b>266</b>	<b>266</b>		<b>263</b>		± 40	<b>485</b>	<b>490</b>		<b>490</b>		± 60
VPM	MPV	µm <sup>3</sup> ; fl	<b>8.3</b>	<b>9.1</b>		<b>9.6</b>		± 2.0	<b>8.4</b>	<b>8.4</b>		<b>8.6</b>		± 2.0	<b>8.5</b>	<b>8.6</b>		<b>8.8</b>		± 2.0
LYMPHO		#	<b>1.4</b>	<b>1.4</b>		N/A		± 0.2	<b>2.9</b>	<b>2.8</b>		N/A		± 0.5	<b>3.8</b>	<b>3.6</b>		N/A		± 1.1
		%	<b>64.5</b>	<b>64.5</b>		N/A		± 8.0	<b>33.0</b>	<b>32.0</b>		N/A		± 6.0	<b>17.5</b>	<b>16.5</b>		N/A		± 5.0
MONO		#	<b>0.2</b>	<b>0.2</b>		N/A		± 0.2	<b>0.5</b>	<b>0.6</b>		N/A		± 0.4	<b>1.1</b>	<b>1.1</b>		N/A		± 1.0
		%	<b>7.5</b>	<b>8.0</b>		N/A		± 7.0	<b>6.0</b>	<b>6.5</b>		N/A		± 5.0	<b>5.0</b>	<b>5.0</b>		N/A		± 5.0
GRANULO		#	<b>0.6</b>	<b>0.6</b>		N/A		± 0.2	<b>5.3</b>	<b>5.4</b>		N/A		± 0.5	<b>16.7</b>	<b>17.1</b>		N/A		± 1.5
		%	<b>28.0</b>	<b>27.5</b>		N/A		± 7.0	<b>61.0</b>	<b>61.5</b>		N/A		± 6.0	<b>77.5</b>	<b>78.5</b>		N/A		± 7.0
CRP		mg/dl	<b>0.58</b>	<b>0.66</b>		<b>0.54</b>		± 0.40	<b>2.39</b>	<b>2.51</b>		<b>2.31</b>		± 0.80	<b>4.42</b>	<b>4.63</b>		<b>4.47</b>		± 1.50
		mg/l	<b>5.8</b>	<b>6.6</b>		<b>5.4</b>		± 4.0	<b>23.9</b>	<b>25.1</b>		<b>23.1</b>		± 8.0	<b>44.2</b>	<b>46.3</b>		<b>44.7</b>		± 15.0

Ref: TEMP-1070 Rev.39 BACKVERSO 1300092170