

# ABX Minotrol CRP



**LOT**

**MC 423**

**CONTROL**



**(Exp.) 2022-07-05**  
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE	
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA				
				EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G			MS CRP		EMI CRP	LC-767G			MS CRP
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.2</b>	<b>2.1</b>		<b>2.2</b>		± 0.4	<b>7.6</b>	<b>7.6</b>		<b>8.0</b>		± 1.0	<b>20.7</b>	<b>21.1</b>		<b>21.6</b>		± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.24</b>	<b>2.23</b>		<b>2.22</b>		± 0.15	<b>4.50</b>	<b>4.50</b>		<b>4.54</b>		± 0.18	<b>5.56</b>	<b>5.62</b>		<b>5.66</b>		± 0.25
HB	HGB	g/dl	<b>6.0</b>	<b>6.0</b>		<b>5.8</b>		± 0.4	<b>13.2</b>	<b>13.3</b>		<b>13.4</b>		± 0.5	<b>18.0</b>	<b>18.0</b>		<b>18.4</b>		± 0.7
		g/l	<b>60</b>	<b>60</b>		<b>58</b>		± 4	<b>132</b>	<b>133</b>		<b>134</b>		± 5	<b>180</b>	<b>180</b>		<b>184</b>		± 7
		mmol/l	<b>3.73</b>	<b>3.73</b>		<b>3.60</b>		± 0.25	<b>8.20</b>	<b>8.26</b>		<b>8.32</b>		± 0.31	<b>11.18</b>	<b>11.18</b>		<b>11.43</b>		± 0.44
HT	HCT	%	<b>16.1</b>	<b>16.5</b>		<b>16.0</b>		± 2.0	<b>36.5</b>	<b>36.5</b>		<b>36.3</b>		± 2.5	<b>49.5</b>	<b>50.0</b>		<b>49.8</b>		± 3.0
		l/l	<b>0.161</b>	<b>0.165</b>		<b>0.160</b>		± 0.020	<b>0.365</b>	<b>0.365</b>		<b>0.363</b>		± 0.025	<b>0.495</b>	<b>0.500</b>		<b>0.498</b>		± 0.030
VGM	MCV	µm <sup>3</sup> ; fl	<b>72</b>	<b>74</b>		<b>72.0</b>		± 4	<b>81</b>	<b>81</b>		<b>80.0</b>		± 4	<b>89</b>	<b>89</b>		<b>88.0</b>		± 4
TGMH	MCH	pg	<b>26.8</b>	<b>26.9</b>		<b>26.1</b>		± 2.4	<b>29.3</b>	<b>29.6</b>		<b>29.5</b>		± 2.4	<b>32.4</b>	<b>32.0</b>		<b>32.5</b>		± 2.8
		fmol	<b>1.66</b>	<b>1.67</b>		<b>1.62</b>		± 0.15	<b>1.82</b>	<b>1.84</b>		<b>1.83</b>		± 0.15	<b>2.01</b>	<b>1.99</b>		<b>2.02</b>		± 0.18
CCMH	MCHC	g/dl	<b>37.2</b>	<b>36.4</b>		<b>36.3</b>		± 3.0	<b>36.2</b>	<b>36.5</b>		<b>36.9</b>		± 3.0	<b>36.4</b>	<b>36.0</b>		<b>36.9</b>		± 3.0
		g/l	<b>372</b>	<b>364</b>		<b>363</b>		± 30	<b>362</b>	<b>365</b>		<b>369</b>		± 30	<b>364</b>	<b>360</b>		<b>369</b>		± 30
		mmol/l	<b>23.10</b>	<b>22.58</b>		<b>22.53</b>		± 1.86	<b>22.49</b>	<b>22.66</b>		<b>22.91</b>		± 1.86	<b>22.59</b>	<b>22.35</b>		<b>22.94</b>		± 1.86
IDR	RDW	%	<b>13.0</b>	<b>13.0</b>		<b>15.0</b>		± 3.0	<b>12.5</b>	<b>12.5</b>		<b>12.5</b>		± 3.0	<b>12.0</b>	<b>12.0</b>		<b>12.0</b>		± 3.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>75</b>	<b>82</b>		<b>68</b>		± 20	<b>254</b>	<b>260</b>		<b>252</b>		± 40	<b>489</b>	<b>498</b>		<b>495</b>		± 60
VPM	MPV	µm <sup>3</sup> ; fl	<b>8.0</b>	<b>8.7</b>		<b>9.4</b>		± 2.0	<b>8.5</b>	<b>8.6</b>		<b>9.0</b>		± 2.0	<b>8.5</b>	<b>8.6</b>		<b>9.0</b>		± 2.0
LYMPHO		#	<b>1.3</b>	<b>1.2</b>		N/A		± 0.2	<b>2.5</b>	<b>2.4</b>		N/A		± 0.5	<b>3.7</b>	<b>3.6</b>		N/A		± 1.1
		%	<b>59.0</b>	<b>59.0</b>		N/A		± 8.0	<b>33.0</b>	<b>32.0</b>		N/A		± 6.0	<b>18.0</b>	<b>17.0</b>		N/A		± 5.0
MONO		#	<b>0.2</b>	<b>0.2</b>		N/A		± 0.2	<b>0.6</b>	<b>0.6</b>		N/A		± 0.4	<b>1.2</b>	<b>1.3</b>		N/A		± 1.0
		%	<b>10.5</b>	<b>11.0</b>		N/A		± 7.0	<b>8.0</b>	<b>8.0</b>		N/A		± 5.0	<b>6.0</b>	<b>6.0</b>		N/A		± 5.0
GRANULO		#	<b>0.7</b>	<b>0.6</b>		N/A		± 0.2	<b>4.5</b>	<b>4.6</b>		N/A		± 0.5	<b>15.7</b>	<b>16.2</b>		N/A		± 1.5
		%	<b>30.5</b>	<b>30.0</b>		N/A		± 7.0	<b>59.0</b>	<b>60.0</b>		N/A		± 6.0	<b>76.0</b>	<b>77.0</b>		N/A		± 7.0
CRP		mg/dl	<b>0.46</b>	<b>0.52</b>		<b>0.57</b>		± 0.30	<b>2.66</b>	<b>2.77</b>		<b>2.69</b>		± 0.70	<b>4.00</b>	<b>4.04</b>		<b>4.23</b>		± 1.00
		mg/l	<b>4.6</b>	<b>5.2</b>		<b>5.7</b>		± 3.0	<b>26.6</b>	<b>27.7</b>		<b>26.9</b>		± 7.0	<b>40.0</b>	<b>40.4</b>		<b>42.3</b>		± 10.0

Ref: TEMP-1070 Rev.39 BACK/VERSO 1300092170