

ABX Minotrol CRP



LOT

MC 422

CONTROL



(Exp.) 2022-04-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE	
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA				
				EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G			MS CRP		EMI CRP	LC-767G			MS CRP
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.1		2.2		± 0.4	8.7	8.7		9.1		± 1.0	21.5	21.8		22.5		± 2.2
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.36	2.33		2.32		± 0.15	4.50	4.47		4.50		± 0.18	5.46	5.40		5.49		± 0.25
HB	HGB	g/dl	6.2	6.2		5.9		± 0.4	13.3	13.4		13.4		± 0.5	17.3	17.4		17.8		± 0.7
		g/l	62	62		59		± 4	133	134		134		± 5	173	174		178		± 7
		mmol/l	3.85	3.85		3.66		± 0.25	8.26	8.32		8.32		± 0.31	10.74	10.81		11.05		± 0.44
HT	HCT	%	16.8	17.0		17.2		± 2.0	36.5	36.7		37.2		± 2.5	48.0	48.1		47.2		± 3.0
		l/l	0.168	0.170		0.172		± 0.020	0.365	0.367		0.372		± 0.025	0.480	0.481		0.472		± 0.030
VGM	MCV	µm ³ ; fl	71	73		74.1		± 4	81	82		82.7		± 4	88	89		86.0		± 4
TGMH	MCH	pg	26.3	26.6		25.4		± 2.4	29.6	30.0		29.8		± 2.4	31.7	32.2		32.4		± 2.8
		fmol	1.63	1.65		1.58		± 0.15	1.84	1.86		1.85		± 0.15	1.97	2.00		2.01		± 0.18
CCMH	MCHC	g/dl	37.0	36.5		34.3		± 3.0	36.5	36.6		36.0		± 3.0	36.0	36.2		37.7		± 3.0
		g/l	370	365		343		± 30	365	366		360		± 30	360	362		377		± 30
		mmol/l	22.98	22.64		21.31		± 1.86	22.66	22.70		22.36		± 1.86	22.36	22.48		23.41		± 1.86
IDR	RDW	%	13.0	13.0		15.0		± 3.0	13.0	13.0		13.0		± 3.0	12.0	12.0		11.5		± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	71	78		67		± 20	266	266		263		± 40	485	490		490		± 60
VPM	MPV	µm ³ ; fl	8.3	9.1		9.6		± 2.0	8.4	8.4		8.6		± 2.0	8.5	8.6		8.8		± 2.0
LYMPHO		#	1.4	1.4		N/A		± 0.2	2.9	2.8		N/A		± 0.5	3.8	3.6		N/A		± 1.1
		%	64.5	64.5		N/A		± 8.0	33.0	32.0		N/A		± 6.0	17.5	16.5		N/A		± 5.0
MONO		#	0.2	0.2		N/A		± 0.2	0.5	0.6		N/A		± 0.4	1.1	1.1		N/A		± 1.0
		%	7.5	8.0		N/A		± 7.0	6.0	6.5		N/A		± 5.0	5.0	5.0		N/A		± 5.0
GRANULO		#	0.6	0.6		N/A		± 0.2	5.3	5.4		N/A		± 0.5	16.7	17.1		N/A		± 1.5
		%	28.0	27.5		N/A		± 7.0	61.0	61.5		N/A		± 6.0	77.5	78.5		N/A		± 7.0
CRP		mg/dl	0.58	0.66		0.54		± 0.40	2.39	2.51		2.31		± 0.80	4.42	4.63		4.47		± 1.50
		mg/l	5.8	6.6		5.4		± 4.0	23.9	25.1		23.1		± 8.0	44.2	46.3		44.7		± 15.0

Ref: TEMP-1070 Rev.39 BACKVERSO 1300092170