

ABX Minotrol CRP



LOT

MC 419

CONTROL



(Exp.) 2021-07-05

(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														
				CONTROL				1	CONTROL				2	CONTROL				3
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
				EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP		
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1		2.1		± 0.4	7.7		8.0		± 1.0	20.5		21.4		± 2.2	
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.42		2.39		± 0.15	4.51		4.57		± 0.18	5.42		5.54		± 0.25	
HB	HGB	g/dl	6.2		5.9		± 0.4	13.2		13.3		± 0.5	17.6		18.1		± 0.7	
		g/l	62		59		± 4	132		133		± 5	176		181		± 7	
		mmol/l	3.85		3.66		± 0.25	8.20		8.26		± 0.31	10.93		11.24		± 0.44	
HT	HCT	%	16.9		16.5		± 2.0	36.1		36.1		± 2.5	48.8		48.8		± 3.0	
		l/l	0.169		0.165		± 0.020	0.361		0.361		± 0.025	0.488		0.488		± 0.030	
VGM	MCV	µm ³ ; fl	70		69.0		± 4	80		79.0		± 4	90		88.0		± 4	
TGMH	MCH	pg	25.6		24.7		± 2.4	29.3		29.1		± 2.4	32.5		32.7		± 2.8	
		fmol	1.59		1.53		± 0.15	1.82		1.81		± 0.15	2.02		2.03		± 0.18	
CCMH	MCHC	g/dl	36.6		35.8		± 3.0	36.6		36.8		± 3.0	36.1		37.1		± 3.0	
		g/l	366		358		± 30	366		368		± 30	361		371		± 30	
		mmol/l	22.73		22.22		± 1.86	22.72		22.88		± 1.86	22.41		23.06		± 1.86	
IDR	RDW	%	13.5		15.5		± 3.0	13.0		13.0		± 3.0	13.0		11.5		± 3.0	
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	64		64		± 20	247		253		± 40	480		510		± 60	
VPM	MPV	µm ³ ; fl	7.8		9.2		± 2.0	8.0		9.0		± 2.0	7.7		8.5		± 2.0	
LYMPHO		#	1.4		N/A		± 0.2	2.7		N/A		± 0.5	3.7		N/A		± 1.1	
		%	66.0		N/A		± 8.0	34.5		N/A		± 6.0	18.0		N/A		± 5.0	
MONO		#	0.2		N/A		± 0.2	0.5		N/A		± 0.4	1.0		N/A		± 1.0	
		%	6.5		N/A		± 6.0	6.5		N/A		± 5.0	5.0		N/A		± 5.0	
GRANULO		#	0.6		N/A		± 0.2	4.5		N/A		± 0.5	15.8		N/A		± 1.5	
		%	27.5		N/A		± 7.0	59.0		N/A		± 6.0	77.0		N/A		± 7.0	
CRP		mg/dl	0.68		0.68		± 0.40	2.23		2.23		± 0.80	4.05		4.08		± 1.50	
		mg/l	6.8		6.8		± 4.0	22.3		22.3		± 8.0	40.5		40.8		± 15.0	

Ref: TEMP-1070 Rev.38 BACK/VERSO 1300060514