

ABX Minotrol CRP



LOT

MC 418

CONTROL



(Exp.) 2021-04-05

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														
				CONTROL				1	CONTROL				2	CONTROL				3
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
				EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP		
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1		2.1		± 0.4	7.7		8.0		± 1.0	20.5		21.9		± 2.2	
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.41		2.37		± 0.15	4.68		4.68		± 0.18	5.55		5.63		± 0.25	
HB	HGB	g/dl	6.2		5.9		± 0.4	13.7		13.8		± 0.5	17.7		18.2		± 0.7	
		g/l	62		59		± 4	137		138		± 5	177		182		± 7	
		mmol/l	3.85		3.66		± 0.25	8.51		8.57		± 0.31	10.99		11.30		± 0.44	
HT	HCT	%	16.9		16.6		± 2.0	37.9		37.4		± 2.5	49.4		49.0		± 3.0	
		l/l	0.169		0.166		± 0.020	0.379		0.374		± 0.025	0.494		0.490		± 0.030	
VGM	MCV	µm ³ ; fl	70		70.0		± 4	81		80.0		± 4	89		87.0		± 4	
TGMH	MCH	pg	25.7		24.9		± 2.4	29.3		29.5		± 2.4	31.9		32.3		± 2.8	
		fmol	1.60		1.55		± 0.15	1.82		1.83		± 0.15	1.98		2.01		± 0.18	
CCMH	MCHC	g/dl	36.8		35.6		± 3.0	36.1		36.9		± 3.0	35.8		37.2		± 3.0	
		g/l	368		356		± 30	361		369		± 30	358		372		± 30	
		mmol/l	22.82		22.08		± 1.86	22.44		22.89		± 1.86	22.25		23.07		± 1.86	
IDR	RDW	%	14.0		15.0		± 3.0	13.0		13.0		± 3.0	12.5		12.5		± 3.0	
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	75		72		± 20	272		272		± 40	519		541		± 60	
VPM	MPV	µm ³ ; fl	8.0		9.6		± 2.0	8.5		9.1		± 2.0	8.3		8.8		± 2.0	
LYMPHO		#	1.3		N/A		± 0.3	2.7		N/A		± 0.5	3.5		N/A		± 1.3	
		%	63.5		N/A		± 8.0	35.5		N/A		± 6.0	17.0		N/A		± 6.0	
MONO		#	0.2		N/A		± 0.2	0.5		N/A		± 0.5	0.9		N/A		± 0.9	
		%	8.0		N/A		± 7.0	6.0		N/A		± 5.0	4.5		N/A		± 4.5	
GRANULO		#	0.6		N/A		± 0.2	4.5		N/A		± 0.6	16.1		N/A		± 1.5	
		%	28.5		N/A		± 7.0	58.5		N/A		± 6.0	78.5		N/A		± 7.0	
CRP		mg/dl	0.53		0.52		± 0.40	2.17		2.17		± 0.80	3.98		4.17		± 1.50	
		mg/l	5.3		5.2		± 4.0	21.7		21.7		± 8.0	39.8		41.7		± 15.0	

Ref: TEMP-1070 Rev.38 BACK/VERSO 1300060514