

ABX Minotrol CRP



LOT

MC 417

CONTROL



(Exp.) 2021-01-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio															
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA			
				EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP			
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0		2.0		± 0.4	7.7		7.9		± 1.0	20.1		20.9		± 2.2		
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.43		2.40		± 0.15	4.55		4.55		± 0.18	5.42		5.45		± 0.25		
HB	HGB	g/dl	6.2		5.9		± 0.4	13.4		13.5		± 0.5	17.6		17.9		± 0.7		
		g/l	62		59		± 4	134		135		± 5	176		179		± 7		
		mmol/l	3.85		3.66		± 0.25	8.32		8.38		± 0.31	10.93		11.12		± 0.44		
HT	HCT	%	17.0		16.6		± 2.0	37.3		36.9		± 2.5	48.8		48.5		± 3.0		
		l/l	0.170		0.166		± 0.020	0.373		0.369		± 0.025	0.488		0.485		± 0.030		
VGM	MCV	µm ³ ; fl	70		69.0		± 4	82		81.0		± 4	90		89.0		± 4		
TGMH	MCH	pg	25.5		24.6		± 2.4	29.5		29.7		± 2.4	32.5		32.8		± 2.8		
		fmol	1.58		1.53		± 0.15	1.83		1.84		± 0.15	2.02		2.04		± 0.18		
CCMH	MCHC	g/dl	36.4		35.6		± 3.0	35.9		36.6		± 3.0	36.1		36.9		± 3.0		
		g/l	364		356		± 30	359		366		± 30	361		369		± 30		
		mmol/l	22.63		22.13		± 1.86	22.30		22.75		± 1.86	22.41		22.92		± 1.86		
IDR	RDW	%	14.0		14.5		± 3.0	13.0		13.2		± 3.0	12.0		11.8		± 3.0		
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	68		67		± 20	251		255		± 40	489		523		± 60		
VPM	MPV	µm ³ ; fl	8.2		9.5		± 2.0	8.4		9.2		± 2.0	8.1		8.4		± 2.0		
LYMPHO		#	1.3		N/A		± 0.2	2.8		N/A		± 0.5	3.5		N/A		± 1.2		
		%	64.5		N/A		± 8.0	36.0		N/A		± 6.0	17.5		N/A		± 6.0		
MONO		#	0.2		N/A		± 0.2	0.5		N/A		± 0.4	0.9		N/A		± 0.9		
		%	8.0		N/A		± 7.0	6.5		N/A		± 5.0	4.5		N/A		± 4.5		
GRANULO		#	0.6		N/A		± 0.2	4.4		N/A		± 0.5	15.7		N/A		± 1.5		
		%	27.5		N/A		± 7.0	57.5		N/A		± 6.0	78.0		N/A		± 7.0		
CRP		mg/dl	0.60		0.56		± 0.40	1.95		1.86		± 0.80	4.10		3.97		± 1.50		
		mg/l	6.0		5.6		± 4.0	19.5		18.6		± 8.0	41.0		39.7		± 15.0		

Ref: TEMP-1070 Rev.38 BACK/VERSO 1300060514