

ABX Minotrol CRP



LOT

MC 416

CONTROL



(Exp.) 2020-10-05
(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio															
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	
				EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP			
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0		2.1		± 0.4	7.8		8.2		± 1.0	20.6		21.6		± 2.2		
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.40		2.37		± 0.15	4.52		4.52		± 0.18	5.55		5.59		± 0.25		
HB	HGB	g/dl	6.1		5.9		± 0.4	13.3		13.5		± 0.5	17.8		18.2		± 0.7		
		g/l	61		59		± 4	133		135		± 5	178		182		± 7		
		mmol/l	3.79		3.66		± 0.25	8.26		8.38		± 0.31	11.05		11.30		± 0.44		
HT	HCT	%	16.6		16.6		± 2.0	36.6		36.6		± 2.5	49.4		48.6		± 3.0		
		l/l	0.166		0.166		± 0.020	0.366		0.366		± 0.025	0.494		0.486		± 0.030		
VGM	MCV	µm ³ ; fl	69		70.0		± 4	81		81.0		± 4	89		87.0		± 4		
TGMH	MCH	pg	25.4		24.9		± 2.4	29.4		29.9		± 2.4	32.1		32.6		± 2.8		
		fmol	1.58		1.55		± 0.15	1.83		1.85		± 0.15	1.99		2.02		± 0.18		
CCMH	MCHC	g/dl	36.8		35.6		± 3.0	36.3		36.9		± 3.0	36.0		37.4		± 3.0		
		g/l	368		356		± 30	363		369		± 30	360		374		± 30		
		mmol/l	22.88		22.08		± 1.86	22.56		22.90		± 1.86	22.38		23.24		± 1.86		
IDR	RDW	%	13.3		14.5		± 3.0	12.8		13.5		± 3.0	12.5		12.5		± 3.0		
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	59		60		± 20	260		266		± 40	496		525		± 60		
VPM	MPV	µm ³ ; fl	7.7		9.0		± 2.0	8.1		8.6		± 2.0	7.9		8.3		± 2.0		
LYMPHO		#	1.3		N/A		± 0.2	2.6		N/A		± 0.5	3.5		N/A		± 1.3		
		%	63.0		N/A		± 8.0	33.0		N/A		± 6.0	17.0		N/A		± 6.0		
MONO		#	0.2		N/A		± 0.2	0.5		N/A		± 0.4	0.9		N/A		± 0.9		
		%	9.5		N/A		± 7.0	7.0		N/A		± 5.0	4.5		N/A		± 4.5		
GRANULO		#	0.6		N/A		± 0.2	4.7		N/A		± 0.5	16.2		N/A		± 1.5		
		%	27.5		N/A		± 7.0	60.0		N/A		± 6.0	78.5		N/A		± 7.0		
CRP		mg/dl	0.54		0.51		± 0.40	2.50		2.14		± 0.80	3.76		3.72		± 1.50		
		mg/l	5.4		5.1		± 4.0	25.0		21.4		± 8.0	37.6		37.2		± 15.0		

Ref: TEMP-1070 Rev.38 BACK/VERSO 1300060514