

ABX Minotrol CRP



LOT

MC 415

CONTROL



(Exp.) 2020-07-05

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio															
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	
				EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP			
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0		2.1		± 0.4	7.7		8.2		± 1.0	20.9		21.9		± 2.2		
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.44		2.47		± 0.15	4.65		4.70		± 0.18	5.57		5.67		± 0.25		
HB	HGB	g/dl	6.3		6.1		± 0.4	13.4		13.6		± 0.5	18.0		18.5		± 0.7		
		g/l	63		61		± 4	134		136		± 5	180		185		± 7		
		mmol/l	3.91		3.79		± 0.25	8.32		8.45		± 0.31	11.18		11.49		± 0.44		
HT	HCT	%	16.8		17.0		± 2.0	37.2		37.1		± 2.5	50.1		49.9		± 3.0		
		l/l	0.168		0.170		± 0.020	0.372		0.371		± 0.025	0.501		0.499		± 0.030		
VGM	MCV	µm ³ ; fl	69		69.0		± 4	80		79.0		± 4	90		88.0		± 4		
TGMH	MCH	pg	25.8		24.7		± 2.4	28.8		28.9		± 2.4	32.3		32.6		± 2.8		
		fmol	1.60		1.53		± 0.15	1.79		1.80		± 0.15	2.01		2.03		± 0.18		
CCMH	MCHC	g/dl	37.4		35.8		± 3.0	36.0		36.6		± 3.0	35.9		37.1		± 3.0		
		g/l	374		358		± 30	360		366		± 30	359		371		± 30		
		mmol/l	23.24		22.23		± 1.86	22.37		22.75		± 1.86	22.30		23.02		± 1.86		
IDR	RDW	%	13.4		14.7		± 3.0	13.1		13.5		± 3.0	12.7		12.1		± 3.0		
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	60		60		± 20	246		250		± 40	495		510		± 60		
VPM	MPV	µm ³ ; fl	8.0		9.2		± 2.0	8.1		8.7		± 2.0	8.2		8.6		± 2.0		
LYMPHO		#	1.3		N/A		± 0.2	2.7		N/A		± 0.5	3.7		N/A		± 1.3		
		%	63.0		N/A		± 8.0	35.0		N/A		± 6.0	17.5		N/A		± 6.0		
MONO		#	0.2		N/A		± 0.2	0.5		N/A		± 0.4	1.0		N/A		± 0.8		
		%	8.0		N/A		± 7.0	6.0		N/A		± 5.0	5.0		N/A		± 4.0		
GRANULO		#	0.6		N/A		± 0.2	4.5		N/A		± 0.5	16.2		N/A		± 1.5		
		%	29.0		N/A		± 7.0	59.0		N/A		± 6.0	77.5		N/A		± 7.0		
CRP		mg/dl	0.68		0.63		± 0.40	2.49		2.23		± 0.80	5.44		4.97		± 1.50		
		mg/l	6.8		6.3		± 4.0	24.9		22.3		± 8.0	54.4		49.7		± 15.0		

Ref: TEMP-1070 Rev.38 BACK/VERSO 1300060514