

ABX Minotrol CRP



LOT

MC 414

CONTROL



(Exp.) 2020-04-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL					1	CONTROL					2	CONTROL					3
		MICROS			PENTRA		TOLERANCES TOLERANCE	MICROS			PENTRA		TOLERANCES TOLERANCE	MICROS			PENTRA		TOLERANCES TOLERANCE
		EMI CRP			MS CRP			EMI CRP			MS CRP			EMI CRP			MS CRP		
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	1.9			1.9		± 0.4	7.9			8.1		± 1.0	20.6			21.5		± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.48			2.41		± 0.15	4.60			4.62		± 0.18	5.60			5.67		± 0.25
HB HGB	g/dl	6.2			6.0		± 0.4	13.4			13.6		± 0.5	17.6			18.1		± 0.7
	g/l	62			60		± 4	134			136		± 5	176			181		± 7
	mmol/l	3.85			3.73		± 0.25	8.32			8.45		± 0.31	10.93			11.24		± 0.44
HT HCT	%	17.1			16.9		± 2.0	37.3			37.0		± 2.5	49.3			48.8		± 3.0
	l/l	0.171			0.169		± 0.020	0.373			0.370		± 0.025	0.493			0.488		± 0.030
VGM MCV	µm ³ ; fl	69			70.0		± 4	81			80.0		± 4	88			86.0		± 4
TGMH MCH	pg	25.0			24.9		± 2.4	29.1			29.4		± 2.4	31.4			31.9		± 2.8
	fmol	1.55			1.55		± 0.15	1.81			1.83		± 0.15	1.95			1.98		± 0.18
CCMH MCHC	g/dl	36.2			35.6		± 3.0	36.0			36.8		± 3.0	35.7			37.1		± 3.0
	g/l	362			356		± 30	360			368		± 30	357			371		± 30
	mmol/l	22.50			22.09		± 1.86	22.33			22.85		± 1.86	22.18			23.05		± 1.86
IDR RDW	%	14.0			14.5		± 3.0	13.5			13.0		± 3.0	13.0			12.5		± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	66			65		± 20	260			264		± 40	555			578		± 60
VPM MPV	µm ³ ; fl	7.9			9.0		± 2.0	8.3			8.8		± 2.0	8.4			8.5		± 2.0
LYMPHO	#	1.2			N/A		± 0.2	2.8			N/A		± 0.5	3.6			N/A		± 1.3
	%	64.5			N/A		± 8.0	35.0			N/A		± 6.0	17.5			N/A		± 6.0
MONO	#	0.2			N/A		± 0.2	0.5			N/A		± 0.4	1.0			N/A		± 0.8
	%	8.0			N/A		± 7.0	6.0			N/A		± 5.0	5.0			N/A		± 4.0
GRANULO	#	0.5			N/A		± 0.2	4.7			N/A		± 0.5	16.0			N/A		± 1.5
	%	27.5			N/A		± 7.0	59.0			N/A		± 6.0	77.5			N/A		± 7.0
CRP	mg/dl	0.62			0.57		± 0.40	2.95			2.60		± 0.80	4.79			4.48		± 1.50
	mg/l	6.2			5.7		± 4.0	29.5			26.0		± 8.0	47.9			44.8		± 15.0

Ref: TEMP-1070 Rev.38 BACK/VERSO 1300060514