

ABX Minotrol CRP



LOT

MC 404

Rev 2

CONTROL



(Exp.) **2017-10-05**

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														
		CONTROL				1	CONTROL				2	CONTROL				3
		MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
		EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP		
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1		2.1		± 0.4	7.7		8.0		± 1.0	20.4		21.5		± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.43		2.44		± 0.15	4.68		4.80		± 0.18	5.63		5.75		± 0.25
HB HGB	g/dl	6.3		6.1		± 0.4	13.3		13.5		± 0.5	17.7		18.3		± 0.7
	g/l	63		61		± 4	133		135		± 5	177		183		± 7
HT HCT	mmol/l	3.91		3.79		± 0.25	8.26		8.38		± 0.31	10.99		11.36		± 0.44
	%	17.0		16.8		± 2.0	37.4		37.7		± 2.5	49.5		49.5		± 3.0
VGM MCV	l/l	0.170		0.168		± 0.020	0.374		0.377		± 0.025	0.495		0.495		± 0.030
	µm ³ ; fl	70		69.0		± 4	80		78.5		± 4	88		86.0		± 4
TGMH MCH	pg	25.9		25.0		± 2.4	28.4		28.1		± 2.4	31.4		31.8		± 2.8
	fmol	1.61		1.55		± 0.15	1.76		1.75		± 0.15	1.95		1.98		± 0.18
CCMH MCHC	g/dl	37.0		36.2		± 3.0	35.5		35.8		± 3.0	35.7		37.0		± 3.0
	g/l	370		362		± 30	355		358		± 30	357		370		± 30
IDR RDW	mmol/l	23.00		22.50		± 1.86	22.06		22.25		± 1.86	22.19		22.98		± 1.86
	%	13.0		14.5		± 3.0	12.5		13.5		± 3.0	12.0		12.0		± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	80		69		± 20	252		250		± 40	475		505		± 60
VPM MPV	µm ³ ; fl	7.8		9.2		± 2.0	8.0		9.2		± 2.0	7.5		8.0		± 2.0
LYMPHO	#	1.3		N/A		± 0.2	2.7		N/A		± 0.5	3.7		N/A		± 1.3
	%	64.0		N/A		± 7.0	35.0		N/A		± 6.0	18.0		N/A		± 6.0
MONO	#	0.2		N/A		± 0.2	0.7		N/A		± 0.4	1.4		N/A		± 0.9
	%	7.5		N/A		± 6.0	9.0		N/A		± 5.0	7.0		N/A		± 4.5
GRANULO	#	0.6		N/A		± 0.2	4.3		N/A		± 0.5	15.3		N/A		± 1.5
	%	28.5		N/A		± 8.0	56.0		N/A		± 6.0	75.0		N/A		± 7.0
CRP	mg/dl	0.68		0.67		± 0.40	2.59		2.35		± 0.80	4.99		4.65		± 1.50
	mg/l	6.8		6.7		± 4.0	25.9		23.5		± 8.0	49.9		46.5		± 15.0

Ref: TEMP-1070 Rev.37 BACKVERSO 9930089-A