

# ABX Minotrol CRP



**LOT**

**MC 403**

**CONTROL**



**(Exp.) 2017-07-05**

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio															
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	
				EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP			
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>1.9</b>		<b>2.0</b>		± 0.4	<b>7.9</b>		<b>8.3</b>		± 1.0	<b>20.6</b>		<b>22.2</b>		± 2.2		
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.28</b>		<b>2.28</b>		± 0.15	<b>4.67</b>		<b>4.70</b>		± 0.18	<b>5.62</b>		<b>5.71</b>		± 0.25		
HB	HGB	g/dl	<b>5.9</b>		<b>5.7</b>		± 0.4	<b>13.4</b>		<b>13.5</b>		± 0.5	<b>17.6</b>		<b>18.0</b>		± 0.7		
		g/l	<b>59</b>		<b>57</b>		± 4	<b>134</b>		<b>135</b>		± 5	<b>176</b>		<b>180</b>		± 7		
		mmol/l	<b>3.66</b>		<b>3.54</b>		± 0.25	<b>8.32</b>		<b>8.38</b>		± 0.31	<b>10.93</b>		<b>11.18</b>		± 0.44		
HT	HCT	%	<b>16.0</b>		<b>16.0</b>		± 2.0	<b>37.8</b>		<b>37.1</b>		± 2.5	<b>49.5</b>		<b>49.1</b>		± 3.0		
		l/l	<b>0.160</b>		<b>0.160</b>		± 0.020	<b>0.378</b>		<b>0.371</b>		± 0.025	<b>0.495</b>		<b>0.491</b>		± 0.030		
VGM	MCV	µm <sup>3</sup> ; fl	<b>70</b>		<b>70.0</b>		± 4	<b>81</b>		<b>79.0</b>		± 4	<b>88</b>		<b>86.0</b>		± 4		
TGMH	MCH	pg	<b>25.9</b>		<b>25.0</b>		± 2.4	<b>28.7</b>		<b>28.7</b>		± 2.4	<b>31.3</b>		<b>31.5</b>		± 2.8		
		fmol	<b>1.61</b>		<b>1.55</b>		± 0.15	<b>1.78</b>		<b>1.78</b>		± 0.15	<b>1.94</b>		<b>1.96</b>		± 0.18		
CCMH	MCHC	g/dl	<b>37.0</b>		<b>35.7</b>		± 3.0	<b>35.4</b>		<b>36.4</b>		± 3.0	<b>35.6</b>		<b>36.7</b>		± 3.0		
		g/l	<b>370</b>		<b>357</b>		± 30	<b>354</b>		<b>364</b>		± 30	<b>356</b>		<b>367</b>		± 30		
		mmol/l	<b>22.96</b>		<b>22.18</b>		± 1.86	<b>22.00</b>		<b>22.58</b>		± 1.86	<b>22.10</b>		<b>22.76</b>		± 1.86		
IDR	RDW	%	<b>13.0</b>		<b>14.0</b>		± 3.0	<b>13.2</b>		<b>13.5</b>		± 3.0	<b>12.5</b>		<b>12.7</b>		± 3.0		
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>75</b>		<b>68</b>		± 20	<b>268</b>		<b>270</b>		± 40	<b>520</b>		<b>530</b>		± 60		
VPM	MPV	µm <sup>3</sup> ; fl	<b>8.4</b>		<b>9.4</b>		± 2.0	<b>8.4</b>		<b>9.0</b>		± 2.0	<b>8.4</b>		<b>8.9</b>		± 2.0		
LYMPHO		#	<b>1.2</b>		N/A		± 0.3	<b>2.8</b>		N/A		± 0.6	<b>3.7</b>		N/A		± 1.3		
		%	<b>62.0</b>		N/A		± 7.0	<b>36.0</b>		N/A		± 6.0	<b>18.0</b>		N/A		± 6.0		
MONO		#	<b>0.2</b>		N/A		± 0.2	<b>0.6</b>		N/A		± 0.4	<b>1.0</b>		N/A		± 1.0		
		%	<b>8.0</b>		N/A		± 6.0	<b>8.0</b>		N/A		± 5.0	<b>5.0</b>		N/A		± 5.0		
GRANULO		#	<b>0.6</b>		N/A		± 0.3	<b>4.4</b>		N/A		± 0.6	<b>15.9</b>		N/A		± 1.5		
		%	<b>30.0</b>		N/A		± 8.0	<b>56.0</b>		N/A		± 6.0	<b>77.0</b>		N/A		± 7.0		
CRP		mg/dl	<b>0.81</b>		<b>0.78</b>		± 0.40	<b>2.52</b>		<b>2.62</b>		± 0.80	<b>3.87</b>		<b>4.18</b>		± 1.50		
		mg/l	<b>8.1</b>		<b>7.8</b>		± 4.0	<b>25.2</b>		<b>26.2</b>		± 8.0	<b>38.7</b>		<b>41.8</b>		± 15.0		

Ref: TEMP-1070 Rev.37 BACKVERSO 9930089-A