

# ABX Minotrol CRP



**LOT**

**MC 412**

**CONTROL**



**(Exp.) 2019-10-05**  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														
				CONTROL				1	CONTROL				2	CONTROL				3
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
				EMI CRP			MS CRP			EMI CRP				MS CRP		EMI CRP		
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>		<b>2.1</b>		± 0.4	<b>7.8</b>		<b>8.2</b>		± 1.0	<b>20.7</b>		<b>22.3</b>		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.35</b>		<b>2.35</b>		± 0.15	<b>4.60</b>		<b>4.65</b>		± 0.18	<b>5.46</b>		<b>5.61</b>		± 0.25	
HB	HGB	g/dl	<b>6.1</b>		<b>6.0</b>		± 0.4	<b>13.2</b>		<b>13.4</b>		± 0.5	<b>17.4</b>		<b>18.1</b>		± 0.7	
		g/l	<b>61</b>		<b>60</b>		± 4	<b>132</b>		<b>134</b>		± 5	<b>174</b>		<b>181</b>		± 7	
		mmol/l	<b>3.79</b>		<b>3.73</b>		± 0.25	<b>8.20</b>		<b>8.32</b>		± 0.31	<b>10.81</b>		<b>11.24</b>		± 0.44	
HT	HCT	%	<b>16.5</b>		<b>16.5</b>		± 2.0	<b>36.3</b>		<b>36.7</b>		± 2.5	<b>48.6</b>		<b>48.8</b>		± 3.0	
		l/l	<b>0.165</b>		<b>0.165</b>		± 0.020	<b>0.363</b>		<b>0.367</b>		± 0.025	<b>0.486</b>		<b>0.488</b>		± 0.030	
VGM	MCV	µm <sup>3</sup> ; fl	<b>70</b>		<b>70.0</b>		± 4	<b>79</b>		<b>79.0</b>		± 4	<b>89</b>		<b>87.0</b>		± 4	
TGMH	MCH	pg	<b>26.0</b>		<b>25.5</b>		± 2.4	<b>28.7</b>		<b>28.8</b>		± 2.4	<b>31.9</b>		<b>32.3</b>		± 2.8	
		fmol	<b>1.61</b>		<b>1.59</b>		± 0.15	<b>1.78</b>		<b>1.79</b>		± 0.15	<b>1.98</b>		<b>2.00</b>		± 0.18	
CCMH	MCHC	g/dl	<b>37.1</b>		<b>36.5</b>		± 3.0	<b>36.3</b>		<b>36.5</b>		± 3.0	<b>35.8</b>		<b>37.1</b>		± 3.0	
		g/l	<b>371</b>		<b>365</b>		± 30	<b>363</b>		<b>365</b>		± 30	<b>358</b>		<b>371</b>		± 30	
		mmol/l	<b>23.03</b>		<b>22.65</b>		± 1.86	<b>22.56</b>		<b>22.65</b>		± 1.86	<b>22.24</b>		<b>23.03</b>		± 1.86	
IDR	RDW	%	<b>13.0</b>		<b>14.5</b>		± 3.0	<b>12.5</b>		<b>13.0</b>		± 3.0	<b>12.0</b>		<b>11.5</b>		± 3.0	
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>70</b>		<b>68</b>		± 20	<b>252</b>		<b>256</b>		± 40	<b>505</b>		<b>522</b>		± 60	
VPM	MPV	µm <sup>3</sup> ; fl	<b>7.9</b>		<b>9.0</b>		± 2.0	<b>8.3</b>		<b>8.8</b>		± 2.0	<b>8.4</b>		<b>8.4</b>		± 2.0	
LYMPHO		#	<b>1.3</b>		N/A		± 0.2	<b>2.8</b>		N/A		± 0.5	<b>3.9</b>		N/A		± 1.3	
		%	<b>66.0</b>		N/A		± 7.0	<b>36.5</b>		N/A		± 6.0	<b>19.0</b>		N/A		± 6.0	
MONO		#	<b>0.2</b>		N/A		± 0.2	<b>0.5</b>		N/A		± 0.4	<b>1.1</b>		N/A		± 0.8	
		%	<b>7.5</b>		N/A		± 6.0	<b>6.5</b>		N/A		± 5.0	<b>5.5</b>		N/A		± 4.0	
GRANULO		#	<b>0.5</b>		N/A		± 0.2	<b>4.4</b>		N/A		± 0.5	<b>15.6</b>		N/A		± 1.5	
		%	<b>26.5</b>		N/A		± 7.0	<b>57.0</b>		N/A		± 6.0	<b>75.5</b>		N/A		± 7.0	
CRP		mg/dl	<b>0.56</b>		<b>0.53</b>		± 0.40	<b>2.34</b>		<b>2.11</b>		± 0.80	<b>4.44</b>		<b>4.09</b>		± 1.50	
		mg/l	<b>5.6</b>		<b>5.3</b>		± 4.0	<b>23.4</b>		<b>21.1</b>		± 8.0	<b>44.4</b>		<b>40.9</b>		± 15.0	

Ref: TEMP-1070 Rev.38 BACK/VERSO 1300060514