

# ABX Minotrol CRP



**LOT**

**MC 402**

**CONTROL**



**(Exp.) 2017-04-05**

(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																			
				CONTROL				1		CONTROL				2		CONTROL				3		TOLERANCES TOLERANCE	
				MICROS		PENTRA		MICROS		PENTRA		MICROS		PENTRA		MICROS		PENTRA					
				EMI CRP		MS CRP		EMI CRP		MS CRP		EMI CRP		MS CRP		EMI CRP		MS CRP					
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>		<b>2.0</b>		± 0.4	<b>7.8</b>		<b>8.1</b>		± 1.0	<b>20.6</b>		<b>21.5</b>		± 2.2						
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.43</b>		<b>2.42</b>		± 0.15	<b>4.67</b>		<b>4.70</b>		± 0.18	<b>5.51</b>		<b>5.61</b>		± 0.25						
HB	HGB	g/dl	<b>6.3</b>		<b>6.0</b>		± 0.4	<b>13.6</b>		<b>13.7</b>		± 0.5	<b>17.5</b>		<b>17.9</b>		± 0.7						
		g/l	<b>63</b>		<b>60</b>		± 4	<b>136</b>		<b>137</b>		± 5	<b>175</b>		<b>179</b>		± 7						
		mmol/l	<b>3.91</b>		<b>3.73</b>		± 0.25	<b>8.45</b>		<b>8.51</b>		± 0.31	<b>10.87</b>		<b>11.12</b>		± 0.44						
HT	HCT	%	<b>17.0</b>		<b>16.9</b>		± 2.0	<b>37.4</b>		<b>37.1</b>		± 2.5	<b>48.5</b>		<b>48.2</b>		± 3.0						
		l/l	<b>0.170</b>		<b>0.169</b>		± 0.020	<b>0.374</b>		<b>0.371</b>		± 0.025	<b>0.485</b>		<b>0.482</b>		± 0.030						
VGM	MCV	µm <sup>3</sup> ; fl	<b>70</b>		<b>70.0</b>		± 4	<b>80</b>		<b>79.0</b>		± 4	<b>88</b>		<b>86.0</b>		± 4						
TGMH	MCH	pg	<b>25.9</b>		<b>24.8</b>		± 2.4	<b>29.1</b>		<b>29.1</b>		± 2.4	<b>31.8</b>		<b>31.9</b>		± 2.8						
		fmol	<b>1.61</b>		<b>1.54</b>		± 0.15	<b>1.81</b>		<b>1.81</b>		± 0.15	<b>1.97</b>		<b>1.98</b>		± 0.18						
CCMH	MCHC	g/dl	<b>37.0</b>		<b>35.4</b>		± 3.0	<b>36.4</b>		<b>36.9</b>		± 3.0	<b>36.1</b>		<b>37.1</b>		± 3.0						
		g/l	<b>370</b>		<b>354</b>		± 30	<b>364</b>		<b>369</b>		± 30	<b>361</b>		<b>371</b>		± 30						
		mmol/l	<b>23.00</b>		<b>22.00</b>		± 1.86	<b>22.61</b>		<b>22.91</b>		± 1.86	<b>22.41</b>		<b>23.04</b>		± 1.86						
IDR	RDW	%	<b>13.0</b>		<b>13.4</b>		± 3.0	<b>12.5</b>		<b>12.5</b>		± 3.0	<b>12.0</b>		<b>12.0</b>		± 3.0						
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>70</b>		<b>65</b>		± 20	<b>250</b>		<b>245</b>		± 40	<b>505</b>		<b>495</b>		± 60						
VPM	MPV	µm <sup>3</sup> ; fl	<b>8.0</b>		<b>9.6</b>		± 2.0	<b>8.6</b>		<b>9.4</b>		± 2.0	<b>8.7</b>		<b>9.3</b>		± 2.0						
LYMPHO		#	<b>1.3</b>		N/A		± 0.3	<b>2.8</b>		N/A		± 0.6	<b>3.8</b>		N/A		± 1.3						
		%	<b>63.0</b>		N/A		± 8.0	<b>36.0</b>		N/A		± 6.0	<b>18.5</b>		N/A		± 6.0						
MONO		#	<b>0.2</b>		N/A		± 0.2	<b>0.5</b>		N/A		± 0.4	<b>1.0</b>		N/A		± 1.0						
		%	<b>9.0</b>		N/A		± 7.0	<b>6.0</b>		N/A		± 5.0	<b>5.0</b>		N/A		± 5.0						
GRANULO		#	<b>0.6</b>		N/A		± 0.3	<b>4.5</b>		N/A		± 0.6	<b>15.8</b>		N/A		± 1.5						
		%	<b>28.0</b>		N/A		± 8.0	<b>58.0</b>		N/A		± 6.0	<b>76.5</b>		N/A		± 7.0						
CRP		mg/dl	<b>0.64</b>		<b>0.67</b>		± 0.40	<b>2.44</b>		<b>2.66</b>		± 0.80	<b>4.45</b>		<b>4.85</b>		± 1.50						
		mg/l	<b>6.4</b>		<b>6.7</b>		± 4.0	<b>24.4</b>		<b>26.6</b>		± 8.0	<b>44.5</b>		<b>48.5</b>		± 15.0						

Ref: TEMP-1070 Rev.37 BACK/VERSO 9930089-A