

# ABX Minotrol CRP



**LOT**

**MC 421**

**CONTROL**



**(Exp.) 2022-01-05**  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio													
				CONTROL			1	CONTROL			2	CONTROL			3	TOLERANCES TOLERANCE	
				MICROS		PENTRA	TOLERANCES TOLERANCE	MICROS		PENTRA	TOLERANCES TOLERANCE	MICROS		PENTRA			
				EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP			
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.2</b>	<b>2.2</b>	<b>2.3</b>	± 0.4	<b>7.7</b>		<b>7.7</b>	<b>7.8</b>	± 1.0		<b>19.9</b>	<b>20.3</b>	<b>20.7</b>	± 2.2		
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.33</b>	<b>2.31</b>	<b>2.29</b>	± 0.15	<b>4.53</b>	<b>4.47</b>	<b>4.48</b>	± 0.18	<b>5.57</b>	<b>5.51</b>	<b>5.60</b>	± 0.25				
HB HGB	g/dl	<b>6.1</b>	<b>6.1</b>	<b>5.9</b>	± 0.4	<b>13.4</b>	<b>13.4</b>	<b>13.4</b>	± 0.5	<b>17.9</b>	<b>17.9</b>	<b>18.3</b>	± 0.7				
	g/l	<b>61</b>	<b>61</b>	<b>59</b>	± 4	<b>134</b>	<b>134</b>	<b>134</b>	± 5	<b>179</b>	<b>179</b>	<b>183</b>	± 7				
	mmol/l	<b>3.79</b>	<b>3.79</b>	<b>3.66</b>	± 0.25	<b>8.32</b>	<b>8.32</b>	<b>8.32</b>	± 0.31	<b>11.12</b>	<b>11.12</b>	<b>11.36</b>	± 0.44				
HT HCT	%	<b>16.5</b>	<b>16.6</b>	<b>16.3</b>	± 2.0	<b>36.7</b>	<b>36.7</b>	<b>36.3</b>	± 2.5	<b>50.1</b>	<b>49.6</b>	<b>49.3</b>	± 3.0				
	l/l	<b>0.165</b>	<b>0.166</b>	<b>0.163</b>	± 0.020	<b>0.367</b>	<b>0.367</b>	<b>0.363</b>	± 0.025	<b>0.501</b>	<b>0.496</b>	<b>0.493</b>	± 0.030				
VGM MCV	µm <sup>3</sup> ; fl	<b>71</b>	<b>72</b>	<b>71.0</b>	± 4	<b>81</b>	<b>82</b>	<b>81.0</b>	± 4	<b>90</b>	<b>90</b>	<b>88.0</b>	± 4				
TGMH MCH	pg	<b>26.2</b>	<b>26.4</b>	<b>25.8</b>	± 2.4	<b>29.6</b>	<b>30.0</b>	<b>29.9</b>	± 2.4	<b>32.1</b>	<b>32.5</b>	<b>32.7</b>	± 2.8				
	fmol	<b>1.63</b>	<b>1.64</b>	<b>1.60</b>	± 0.15	<b>1.84</b>	<b>1.86</b>	<b>1.86</b>	± 0.15	<b>2.00</b>	<b>2.02</b>	<b>2.03</b>	± 0.18				
CCMH MCHC	g/dl	<b>36.9</b>	<b>36.7</b>	<b>36.3</b>	± 3.0	<b>36.5</b>	<b>36.6</b>	<b>36.9</b>	± 3.0	<b>35.7</b>	<b>36.1</b>	<b>37.1</b>	± 3.0				
	g/l	<b>369</b>	<b>367</b>	<b>363</b>	± 30	<b>365</b>	<b>366</b>	<b>369</b>	± 30	<b>357</b>	<b>361</b>	<b>371</b>	± 30				
	mmol/l	<b>22.90</b>	<b>22.78</b>	<b>22.53</b>	± 1.86	<b>22.68</b>	<b>22.70</b>	<b>22.93</b>	± 1.86	<b>22.17</b>	<b>22.42</b>	<b>23.06</b>	± 1.86				
IDR RDW	%	<b>13.5</b>	<b>13.0</b>	<b>14.0</b>	± 3.0	<b>13.5</b>	<b>13.0</b>	<b>13.5</b>	± 3.0	<b>12.5</b>	<b>13.0</b>	<b>12.0</b>	± 3.0				
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>68</b>	<b>71</b>	<b>64</b>	± 20	<b>257</b>	<b>263</b>	<b>247</b>	± 40	<b>486</b>	<b>498</b>	<b>499</b>	± 60				
VPM MPV	µm <sup>3</sup> ; fl	<b>8.4</b>	<b>9.1</b>	<b>9.7</b>	± 2.0	<b>8.5</b>	<b>8.5</b>	<b>9.1</b>	± 2.0	<b>8.3</b>	<b>8.5</b>	<b>8.6</b>	± 2.0				
LYMPHO	#	<b>1.4</b>	<b>1.3</b>	N/A	± 0.2	<b>2.5</b>	<b>2.4</b>	N/A	± 0.5	<b>3.1</b>	<b>3.0</b>	N/A	± 1.1				
	%	<b>62.0</b>	<b>61.0</b>	N/A	± 8.0	<b>32.0</b>	<b>31.5</b>	N/A	± 6.0	<b>15.5</b>	<b>15.0</b>	N/A	± 5.0				
MONO	#	<b>0.2</b>	<b>0.2</b>	N/A	± 0.2	<b>0.6</b>	<b>0.6</b>	N/A	± 0.4	<b>1.1</b>	<b>1.1</b>	N/A	± 1.0				
	%	<b>8.5</b>	<b>9.5</b>	N/A	± 7.0	<b>7.5</b>	<b>7.5</b>	N/A	± 5.0	<b>5.5</b>	<b>5.5</b>	N/A	± 5.0				
GRANULO	#	<b>0.6</b>	<b>0.6</b>	N/A	± 0.2	<b>4.7</b>	<b>4.7</b>	N/A	± 0.5	<b>15.7</b>	<b>16.1</b>	N/A	± 1.5				
	%	<b>29.5</b>	<b>29.5</b>	N/A	± 7.0	<b>60.5</b>	<b>61.0</b>	N/A	± 6.0	<b>79.0</b>	<b>79.5</b>	N/A	± 7.0				
CRP	mg/dl	<b>0.56</b>	<b>0.68</b>	<b>0.58</b>	± 0.40	<b>2.03</b>	<b>2.35</b>	<b>2.08</b>	± 0.80	<b>3.14</b>	<b>3.51</b>	<b>3.27</b>	± 1.50				
	mg/l	<b>5.6</b>	<b>6.8</b>	<b>5.8</b>	± 4.0	<b>20.3</b>	<b>23.5</b>	<b>20.8</b>	± 8.0	<b>31.4</b>	<b>35.1</b>	<b>32.7</b>	± 15.0				

Ref: TEMP-1070 Rev.39 BACK/VERSO 1300092170