

# ABX Minotrol CRP



**LOT**

**MC 411**

**CONTROL**



**(Exp.) 2019-07-05**

(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														
				CONTROL		1	CONTROL		2	CONTROL		3	TOLERANCES TOLERANCE					
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
				EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP		
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.2</b>		<b>2.3</b>		± 0.4	<b>7.7</b>		<b>8.1</b>		± 1.0	<b>20.0</b>		<b>21.6</b>		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.44</b>		<b>2.39</b>		± 0.15	<b>4.56</b>		<b>4.57</b>		± 0.18	<b>5.54</b>		<b>5.62</b>		± 0.25	
HB	HGB	g/dl	<b>6.1</b>		<b>5.9</b>		± 0.4	<b>13.3</b>		<b>13.5</b>		± 0.5	<b>17.7</b>		<b>18.2</b>		± 0.7	
		g/l	<b>61</b>		<b>59</b>		± 4	<b>133</b>		<b>135</b>		± 5	<b>177</b>		<b>182</b>		± 7	
		mmol/l	<b>3.79</b>		<b>3.66</b>		± 0.25	<b>8.26</b>		<b>8.38</b>		± 0.31	<b>10.99</b>		<b>11.30</b>		± 0.44	
HT	HCT	%	<b>16.6</b>		<b>16.5</b>		± 2.0	<b>36.5</b>		<b>36.3</b>		± 2.5	<b>49.3</b>		<b>48.3</b>		± 3.0	
		l/l	<b>0.166</b>		<b>0.165</b>		± 0.020	<b>0.365</b>		<b>0.363</b>		± 0.025	<b>0.493</b>		<b>0.483</b>		± 0.030	
VGM	MCV	µm <sup>3</sup> ; fl	<b>68</b>		<b>69.0</b>		± 4	<b>80</b>		<b>79.5</b>		± 4	<b>89</b>		<b>86.0</b>		± 4	
TGMH	MCH	pg	<b>25.0</b>		<b>24.7</b>		± 2.4	<b>29.2</b>		<b>29.5</b>		± 2.4	<b>31.9</b>		<b>32.4</b>		± 2.8	
		fmol	<b>1.55</b>		<b>1.53</b>		± 0.15	<b>1.81</b>		<b>1.83</b>		± 0.15	<b>1.98</b>		<b>2.01</b>		± 0.18	
CCMH	MCHC	g/dl	<b>36.8</b>		<b>35.8</b>		± 3.0	<b>36.5</b>		<b>37.2</b>		± 3.0	<b>35.9</b>		<b>37.7</b>		± 3.0	
		g/l	<b>368</b>		<b>358</b>		± 30	<b>365</b>		<b>372</b>		± 30	<b>359</b>		<b>377</b>		± 30	
		mmol/l	<b>22.83</b>		<b>22.22</b>		± 1.86	<b>22.64</b>		<b>23.08</b>		± 1.86	<b>22.29</b>		<b>23.38</b>		± 1.86	
IDR	RDW	%	<b>13.0</b>		<b>14.5</b>		± 3.0	<b>12.8</b>		<b>13.0</b>		± 3.0	<b>11.7</b>		<b>12.0</b>		± 3.0	
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>69</b>		<b>65</b>		± 20	<b>265</b>		<b>266</b>		± 40	<b>505</b>		<b>517</b>		± 60	
VPM	MPV	µm <sup>3</sup> ; fl	<b>8.8</b>		<b>9.6</b>		± 2.0	<b>8.6</b>		<b>8.8</b>		± 2.0	<b>8.6</b>		<b>8.8</b>		± 2.0	
LYMPHO		#	<b>1.4</b>		N/A		± 0.2	<b>2.6</b>		N/A		± 0.5	<b>3.2</b>		N/A		± 1.2	
		%	<b>64.0</b>		N/A		± 7.0	<b>34.0</b>		N/A		± 6.0	<b>16.0</b>		N/A		± 6.0	
MONO		#	<b>0.2</b>		N/A		± 0.2	<b>0.5</b>		N/A		± 0.4	<b>0.8</b>		N/A		± 0.8	
		%	<b>6.5</b>		N/A		± 6.0	<b>6.0</b>		N/A		± 5.0	<b>4.0</b>		N/A		± 4.0	
GRANULO		#	<b>0.6</b>		N/A		± 0.2	<b>4.6</b>		N/A		± 0.5	<b>16.0</b>		N/A		± 1.4	
		%	<b>29.5</b>		N/A		± 7.0	<b>60.0</b>		N/A		± 6.0	<b>80.0</b>		N/A		± 7.0	
CRP		mg/dl	<b>0.68</b>		<b>0.61</b>		± 0.40	<b>2.76</b>		<b>2.34</b>		± 0.80	<b>4.80</b>		<b>4.10</b>		± 1.50	
		mg/l	<b>6.8</b>		<b>6.1</b>		± 4.0	<b>27.6</b>		<b>23.4</b>		± 8.0	<b>48.0</b>		<b>41.0</b>		± 15.0	

Ref: TEMP-1070 Rev.38 BACKVERSO 1300060514