

# ABX Minotrol CRP



**LOT**

**MC 410**

**CONTROL**



**(Exp.) 2019-04-05**

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														
				CONTROL				1	CONTROL				2	CONTROL				3
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
				EMI CRP			MS CRP			EMI CRP				MS CRP		EMI CRP		
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>		<b>2.1</b>		± 0.4	<b>7.8</b>		<b>8.0</b>		± 1.0	<b>20.0</b>		<b>21.4</b>		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.40</b>		<b>2.38</b>		± 0.15	<b>4.70</b>		<b>4.74</b>		± 0.18	<b>5.59</b>		<b>5.75</b>		± 0.25	
HB	HGB	g/dl	<b>6.2</b>		<b>6.0</b>		± 0.4	<b>13.5</b>		<b>13.6</b>		± 0.5	<b>17.8</b>		<b>18.3</b>		± 0.7	
		g/l	<b>62</b>		<b>60</b>		± 4	<b>135</b>		<b>136</b>		± 5	<b>178</b>		<b>183</b>		± 7	
		mmol/l	<b>3.85</b>		<b>3.73</b>		± 0.25	<b>8.38</b>		<b>8.45</b>		± 0.31	<b>11.05</b>		<b>11.36</b>		± 0.44	
HT	HCT	%	<b>16.8</b>		<b>16.7</b>		± 2.0	<b>37.1</b>		<b>37.2</b>		± 2.5	<b>49.2</b>		<b>49.2</b>		± 3.0	
		l/l	<b>0.168</b>		<b>0.167</b>		± 0.020	<b>0.371</b>		<b>0.372</b>		± 0.025	<b>0.492</b>		<b>0.492</b>		± 0.030	
VGM	MCV	µm <sup>3</sup> ; fl	<b>70</b>		<b>70.0</b>		± 4	<b>79</b>		<b>78.5</b>		± 4	<b>88</b>		<b>85.5</b>		± 4	
TGMH	MCH	pg	<b>25.8</b>		<b>25.2</b>		± 2.4	<b>28.7</b>		<b>28.7</b>		± 2.4	<b>31.8</b>		<b>31.8</b>		± 2.8	
		fmol	<b>1.60</b>		<b>1.57</b>		± 0.15	<b>1.78</b>		<b>1.78</b>		± 0.15	<b>1.98</b>		<b>1.98</b>		± 0.18	
CCMH	MCHC	g/dl	<b>36.9</b>		<b>36.0</b>		± 3.0	<b>36.4</b>		<b>36.6</b>		± 3.0	<b>36.2</b>		<b>37.2</b>		± 3.0	
		g/l	<b>369</b>		<b>360</b>		± 30	<b>364</b>		<b>366</b>		± 30	<b>362</b>		<b>372</b>		± 30	
		mmol/l	<b>22.92</b>		<b>22.36</b>		± 1.86	<b>22.58</b>		<b>22.70</b>		± 1.86	<b>22.47</b>		<b>23.12</b>		± 1.86	
IDR	RDW	%	<b>13.0</b>		<b>14.0</b>		± 3.0	<b>12.5</b>		<b>13.2</b>		± 3.0	<b>12.0</b>		<b>12.0</b>		± 3.0	
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>73</b>		<b>70</b>		± 20	<b>265</b>		<b>260</b>		± 40	<b>500</b>		<b>525</b>		± 60	
VPM	MPV	µm <sup>3</sup> ; fl	<b>7.5</b>		<b>8.5</b>		± 2.0	<b>8.5</b>		<b>8.8</b>		± 2.0	<b>8.3</b>		<b>8.5</b>		± 2.0	
LYMPHO	#		<b>1.3</b>		N/A		± 0.2	<b>2.7</b>		N/A		± 0.5	<b>3.3</b>		N/A		± 1.2	
	%		<b>63.0</b>		N/A		± 8.0	<b>35.0</b>		N/A		± 6.0	<b>16.5</b>		N/A		± 6.0	
MONO	#		<b>0.2</b>		N/A		± 0.2	<b>0.5</b>		N/A		± 0.4	<b>1.0</b>		N/A		± 1.0	
	%		<b>9.0</b>		N/A		± 7.0	<b>7.0</b>		N/A		± 5.0	<b>5.0</b>		N/A		± 5.0	
GRANULO	#		<b>0.6</b>		N/A		± 0.2	<b>4.5</b>		N/A		± 0.5	<b>15.7</b>		N/A		± 1.4	
	%		<b>28.0</b>		N/A		± 7.0	<b>58.0</b>		N/A		± 6.0	<b>78.5</b>		N/A		± 7.0	
CRP	mg/dl		<b>0.66</b>		<b>0.65</b>		± 0.40	<b>2.25</b>		<b>2.05</b>		± 0.80	<b>4.00</b>		<b>3.79</b>		± 1.50	
	mg/l		<b>6.6</b>		<b>6.5</b>		± 4.0	<b>22.5</b>		<b>20.5</b>		± 8.0	<b>40.0</b>		<b>37.9</b>		± 15.0	

Ref: TEMP-1070 Rev.38 BACKVERSO 1300060514