

# ABX Minotrol CRP



**LOT**

**MC 408**

**CONTROL**



**(Exp.) 2018-10-05**

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio													
				CONTROL 1				CONTROL 2				CONTROL 3				TOLERANCES TOLERANCE	
				MICROS		PENTRA		MICROS		PENTRA		MICROS		PENTRA			
				EMI CRP		MS CRP		EMI CRP		MS CRP		EMI CRP		MS CRP			
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>		<b>2.2</b>		± 0.4	<b>7.7</b>		<b>8.1</b>		± 1.0	<b>20.9</b>		<b>22.2</b>		± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.33</b>		<b>2.30</b>		± 0.15	<b>4.55</b>		<b>4.63</b>		± 0.18	<b>5.51</b>		<b>5.65</b>		± 0.25
HB	HGB	g/dl	<b>6.1</b>		<b>5.9</b>		± 0.4	<b>13.3</b>		<b>13.6</b>		± 0.5	<b>17.7</b>		<b>18.3</b>		± 0.7
		g/l	<b>61</b>		<b>59</b>		± 4	<b>133</b>		<b>136</b>		± 5	<b>177</b>		<b>183</b>		± 7
		mmol/l	<b>3.79</b>		<b>3.66</b>		± 0.25	<b>8.26</b>		<b>8.45</b>		± 0.31	<b>10.99</b>		<b>11.36</b>		± 0.44
HT	HCT	%	<b>16.5</b>		<b>16.1</b>		± 2.0	<b>36.4</b>		<b>36.6</b>		± 2.5	<b>49.0</b>		<b>48.6</b>		± 3.0
		l/l	<b>0.165</b>		<b>0.161</b>		± 0.020	<b>0.364</b>		<b>0.366</b>		± 0.025	<b>0.490</b>		<b>0.486</b>		± 0.030
VGM	MCV	µm <sup>3</sup> ; fl	<b>71</b>		<b>70.0</b>		± 4	<b>80</b>		<b>79.0</b>		± 4	<b>89</b>		<b>86.0</b>		± 4
TGMH	MCH	pg	<b>26.2</b>		<b>25.7</b>		± 2.4	<b>29.2</b>		<b>29.4</b>		± 2.4	<b>32.1</b>		<b>32.4</b>		± 2.8
		fmol	<b>1.63</b>		<b>1.59</b>		± 0.15	<b>1.82</b>		<b>1.82</b>		± 0.15	<b>1.99</b>		<b>2.01</b>		± 0.18
CCMH	MCHC	g/dl	<b>36.9</b>		<b>36.6</b>		± 3.0	<b>36.5</b>		<b>37.2</b>		± 3.0	<b>36.1</b>		<b>37.7</b>		± 3.0
		g/l	<b>369</b>		<b>366</b>		± 30	<b>365</b>		<b>372</b>		± 30	<b>361</b>		<b>377</b>		± 30
		mmol/l	<b>22.90</b>		<b>22.76</b>		± 1.86	<b>22.69</b>		<b>23.09</b>		± 1.86	<b>22.41</b>		<b>23.39</b>		± 1.86
IDR	RDW	%	<b>12.6</b>		<b>14.0</b>		± 3.0	<b>13.0</b>		<b>13.0</b>		± 3.0	<b>11.8</b>		<b>11.8</b>		± 3.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>82</b>		<b>71</b>		± 20	<b>255</b>		<b>265</b>		± 40	<b>505</b>		<b>518</b>		± 60
VPM	MPV	µm <sup>3</sup> ; fl	<b>7.4</b>		<b>9.1</b>		± 2.0	<b>7.7</b>		<b>8.6</b>		± 2.0	<b>8.0</b>		<b>8.6</b>		± 2.0
LYMPHO		#	<b>1.4</b>		N/A		± 0.2	<b>2.7</b>		N/A		± 0.5	<b>3.4</b>		N/A		± 1.3
		%	<b>65.0</b>		N/A		± 8.0	<b>35.5</b>		N/A		± 6.0	<b>16.5</b>		N/A		± 6.0
MONO		#	<b>0.2</b>		N/A		± 0.2	<b>0.5</b>		N/A		± 0.4	<b>1.1</b>		N/A		± 1.0
		%	<b>6.5</b>		N/A		± 6.5	<b>6.5</b>		N/A		± 5.0	<b>5.5</b>		N/A		± 5.0
GRANULO		#	<b>0.6</b>		N/A		± 0.2	<b>4.5</b>		N/A		± 0.5	<b>16.3</b>		N/A		± 1.5
		%	<b>28.5</b>		N/A		± 7.0	<b>58.0</b>		N/A		± 6.0	<b>78.0</b>		N/A		± 7.0
CRP		mg/dl	<b>0.59</b>		<b>0.56</b>		± 0.40	<b>2.60</b>		<b>2.37</b>		± 0.80	<b>4.76</b>		<b>4.62</b>		± 1.50
		mg/l	<b>5.9</b>		<b>5.6</b>		± 4.0	<b>26.0</b>		<b>23.7</b>		± 8.0	<b>47.6</b>		<b>46.2</b>		± 15.0

Ref: TEMP-1070 Rev.37 BACKVERSO 9930089-A