

ABX Minotrol CRP



LOT

MC 046

CONTROL



(Exp.) 2016-07-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio												TOLERANCES TOLERANCE							
				CONTROL			1			CONTROL			2					CONTROL			3		
				MICROS		PENTRA	MICROS		PENTRA	MICROS		PENTRA	MICROS		PENTRA			MICROS		PENTRA			
				EMI CRP		MS CRP	EMI CRP		MS CRP	EMI CRP		MS CRP	EMI CRP		MS CRP			EMI CRP		MS CRP			
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1		2.1		± 0.4	7.6		7.9		± 1.0	20.1		21.3		± 2.2						
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.31		2.33		± 0.15	4.55		4.61		± 0.18	5.49		5.59		± 0.25						
HB	HGB	g/dl	6.2		6.0		± 0.4	13.3		13.4		± 0.5	17.5		17.9		± 0.7						
		g/l	62		60		± 4	133		134		± 5	175		179		± 7						
		mmol/l	3.85		3.73		± 0.25	8.26		8.32		± 0.31	10.87		11.12		± 0.44						
HT	HCT	%	16.4		16.5		± 2.0	36.9		36.9		± 2.5	48.9		48.6		± 3.0						
		l/l	0.164		0.165		± 0.020	0.369		0.369		± 0.025	0.489		0.486		± 0.030						
VGM	MCV	µm ³ ; fl	71		71.0		± 4	81		80.0		± 4	89		87.0		± 4						
TGMH	MCH	pg	26.8		25.8		± 2.4	29.2		29.1		± 2.4	31.9		32.0		± 2.8						
		fmol	1.67		1.60		± 0.15	1.82		1.81		± 0.15	1.98		1.99		± 0.18						
CCMH	MCHC	g/dl	37.8		36.3		± 3.0	36.1		36.3		± 3.0	35.8		36.8		± 3.0						
		g/l	378		363		± 30	361		363		± 30	358		368		± 30						
		mmol/l	23.48		22.52		± 1.86	22.41		22.56		± 1.86	22.24		22.86		± 1.86						
IDR	RDW	%	13.0		12.5		± 3.0	12.7		13.0		± 3.0	12.3		11.6		± 3.0						
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	70		60		± 20	265		257		± 40	510		510		± 60						
VPM	MPV	µm ³ ; fl	8.7		9.5		± 2.0	8.4		9.0		± 2.0	8.3		8.4		± 2.0						
LYMPHO		#	1.3		N/A		± 0.4	2.7		N/A		± 0.5	3.6		N/A		± 1.3						
		%	63.0		N/A		± 8.0	36.0		N/A		± 6.0	18.0		N/A		± 6.0						
MONO		#	0.2		N/A		± 0.2	0.5		N/A		± 0.4	1.0		N/A		± 1.0						
		%	9.5		N/A		± 7.0	6.5		N/A		± 5.0	5.0		N/A		± 5.0						
GRANULO		#	0.6		N/A		± 0.3	4.4		N/A		± 0.6	15.5		N/A		± 1.5						
		%	27.5		N/A		± 8.0	57.5		N/A		± 6.0	77.0		N/A		± 7.0						
CRP		mg/dl	0.70		0.70		± 0.40	2.70		2.60		± 0.80	4.30		4.50		± 1.50						
		mg/l	7.0		7.0		± 4.0	27.0		26.0		± 8.0	43.0		45.0		± 15.0						

Ref: TEMP-1070 Rev.36 BACK/VERSO 9930089-A