

ABX Minotrol CRP



LOT

MC 016

CONTROL



(Exp.) 2016-04-05
(YYYY - MM - DD)

Rev 3

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio												TOLERANCES TOLERANCE	
				CONTROL		1	CONTROL		2	CONTROL		3					
				MICROS		PENTRA	MICROS		PENTRA	MICROS		PENTRA					
				EMI CRP		MS CRP	EMI CRP		MS CRP	EMI CRP		MS CRP					
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0		2.1		± 0.4	7.5		7.8		± 1.0	19.7		21.2		± 2.2
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.44		2.45		± 0.15	4.58		4.66		± 0.18	5.54		5.61		± 0.25
HB	HGB	g/dl	6.0		5.9		± 0.4	13.3		13.4		± 0.5	17.9		18.1		± 0.7
		g/l	60		59		± 4	133		134		± 5	179		181		± 7
		mmol/l	3.73		3.66		± 0.25	8.26		8.32		± 0.31	11.12		11.24		± 0.44
HT	HCT	%	16.6		16.7		± 2.0	36.6		36.3		± 2.5	49.3		48.8		± 3.0
		l/l	0.166		0.167		± 0.020	0.366		0.363		± 0.025	0.493		0.488		± 0.030
VGM	MCV	µm ³ ; fl	68		68.0		± 4	80		78.0		± 4	89		87.0		± 4
TGMH	MCH	pg	24.6		24.1		± 2.4	29.0		28.8		± 2.4	32.3		32.3		± 2.8
		fmol	1.53		1.50		± 0.15	1.80		1.79		± 0.15	2.01		2.00		± 0.18
CCMH	MCHC	g/dl	36.2		35.4		± 3.0	36.3		36.9		± 3.0	36.3		37.1		± 3.0
		g/l	362		354		± 30	363		369		± 30	363		371		± 30
		mmol/l	22.46		21.99		± 1.86	22.54		22.89		± 1.86	22.54		23.03		± 1.86
IDR	RDW	%	13.5		13.0		± 3.0	12.5		12.0		± 3.0	12.0		11.0		± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	75		70		± 20	250		248		± 40	490		495		± 60
VPM	MPV	µm ³ ; fl	8.5		9.5		± 2.0	8.0		8.8		± 2.0	8.2		8.6		± 2.0
LYMPHO		#	1.3		N/A		± 0.4	2.6		N/A		± 0.5	3.3		N/A		± 1.2
		%	64.0		N/A		± 8.0	35.0		N/A		± 6.0	16.5		N/A		± 6.0
MONO		#	0.2		N/A		± 0.2	0.6		N/A		± 0.4	1.1		N/A		± 1.0
		%	9.5		N/A		± 7.0	7.5		N/A		± 5.0	5.5		N/A		± 5.0
GRANULO		#	0.5		N/A		± 0.4	4.3		N/A		± 0.5	15.3		N/A		± 1.4
		%	26.5		N/A		± 8.0	57.5		N/A		± 6.0	78.0		N/A		± 7.0
CRP		mg/dl	0.73		0.73		± 0.40	2.40		2.35		± 0.80	3.80		3.90		± 1.50
		mg/l	7.3		7.3		± 4.0	24.0		23.5		± 8.0	38.0		39.0		± 15.0

Ref: TEMP-1070 Rev.36 BACK/VERSO 9930089-A