

# ABX Minotrol CRP



**LOT**

**MC 105**

Rev 2

**CONTROL**



(Exp.) **2016-01-05**

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														
		CONTROL				1	CONTROL				2	CONTROL				3
		MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
		EMI CRP		MS CRP			EMI CRP		MS CRP			EMI CRP		MS CRP		
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>		<b>2.2</b>		± 0.4	<b>8.1</b>		<b>8.1</b>		± 1.0	<b>20.3</b>		<b>20.9</b>		± 2.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.44</b>		<b>2.40</b>		± 0.15	<b>4.52</b>		<b>4.50</b>		± 0.18	<b>5.56</b>		<b>5.55</b>		± 0.25
HB HGB	g/dl	<b>6.1</b>		<b>6.0</b>		± 0.4	<b>13.4</b>		<b>13.4</b>		± 0.5	<b>17.6</b>		<b>17.9</b>		± 0.7
	g/l	<b>61</b>		<b>60</b>		± 4	<b>134</b>		<b>134</b>		± 5	<b>176</b>		<b>179</b>		± 7
HT HCT	mmol/l	<b>3.79</b>		<b>3.73</b>		± 0.25	<b>8.32</b>		<b>8.32</b>		± 0.31	<b>10.93</b>		<b>11.12</b>		± 0.44
	%	<b>17.1</b>		<b>16.7</b>		± 2.0	<b>37.1</b>		<b>36.0</b>		± 2.5	<b>49.5</b>		<b>48.3</b>		± 3.0
VGM MCV	l/l	<b>0.171</b>		<b>0.167</b>		± 0.020	<b>0.371</b>		<b>0.360</b>		± 0.025	<b>0.495</b>		<b>0.483</b>		± 0.030
	µm <sup>3</sup> ; fl	<b>70</b>		<b>69.5</b>		± 4	<b>82</b>		<b>80.0</b>		± 4	<b>89</b>		<b>87.0</b>		± 4
TGMH MCH	pg	<b>25.0</b>		<b>25.0</b>		± 2.4	<b>29.6</b>		<b>29.8</b>		± 2.4	<b>31.7</b>		<b>32.3</b>		± 2.8
	fmol	<b>1.55</b>		<b>1.55</b>		± 0.15	<b>1.84</b>		<b>1.85</b>		± 0.15	<b>1.97</b>		<b>2.00</b>		± 0.18
CCMH MCHC	g/dl	<b>35.7</b>		<b>36.0</b>		± 3.0	<b>36.2</b>		<b>37.2</b>		± 3.0	<b>35.6</b>		<b>37.1</b>		± 3.0
	g/l	<b>357</b>		<b>360</b>		± 30	<b>362</b>		<b>372</b>		± 30	<b>356</b>		<b>371</b>		± 30
IDR RDW	mmol/l	<b>22.18</b>		<b>22.34</b>		± 1.86	<b>22.45</b>		<b>23.12</b>		± 1.86	<b>22.09</b>		<b>23.02</b>		± 1.86
	%	<b>13.1</b>		<b>12.2</b>		± 3.0	<b>12.6</b>		<b>11.2</b>		± 3.0	<b>12.1</b>		<b>11.5</b>		± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>72</b>		<b>58</b>		± 20	<b>260</b>		<b>245</b>		± 40	<b>485</b>		<b>480</b>		± 60
VPM MPV	µm <sup>3</sup> ; fl	<b>8.7</b>		<b>9.8</b>		± 2.0	<b>8.3</b>		<b>9.3</b>		± 2.0	<b>8.2</b>		<b>8.8</b>		± 2.0
LYMPHO	#	<b>1.3</b>		N/A		± 0.3	<b>2.9</b>		N/A		± 0.5	<b>3.7</b>		N/A		± 1.3
	%	<b>63.0</b>		N/A		± 8.0	<b>35.5</b>		N/A		± 6.0	<b>18.0</b>		N/A		± 6.0
MONO	#	<b>0.2</b>		N/A		± 0.2	<b>0.6</b>		N/A		± 0.5	<b>1.0</b>		N/A		± 1.0
	%	<b>10.0</b>		N/A		± 7.0	<b>8.0</b>		N/A		± 5.0	<b>5.0</b>		N/A		± 5.0
GRANULO	#	<b>0.6</b>		N/A		± 0.2	<b>4.6</b>		N/A		± 0.6	<b>15.6</b>		N/A		± 1.5
	%	<b>27.0</b>		N/A		± 7.0	<b>56.5</b>		N/A		± 6.0	<b>77.0</b>		N/A		± 7.0
CRP	mg/dl	<b>0.55</b>		<b>0.50</b>		± 0.40	<b>2.20</b>		<b>2.20</b>		± 0.80	<b>3.60</b>		<b>3.60</b>		± 1.50
	mg/l	<b>5.5</b>		<b>5.0</b>		± 4.0	<b>22.0</b>		<b>22.0</b>		± 8.0	<b>36.0</b>		<b>36.0</b>		± 15.0

Ref: TEMP-1070 Rev.36 BACKVERSO 9930089-A