

# ABX Minotrol CRP



**LOT**

**MC 045**

**CONTROL**



**(Exp.) 2015-07-05**  
(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio												TOLERANCES TOLERANCE													
				CONTROL				1				CONTROL						2				CONTROL				3			
				MICROS		PENTRA		MICROS		PENTRA		MICROS		PENTRA				MICROS		PENTRA		MICROS		PENTRA					
				EMI CRP		MS CRP		EMI CRP		MS CRP		EMI CRP		MS CRP				EMI CRP		MS CRP		EMI CRP		MS CRP					
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>		<b>2.0</b>		± 0.4	<b>7.9</b>		<b>7.9</b>		± 1.0	<b>20.6</b>		<b>21.0</b>		± 2.2												
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.57</b>		<b>2.52</b>		± 0.15	<b>4.61</b>		<b>4.60</b>		± 0.18	<b>5.75</b>		<b>5.73</b>		± 0.25												
HB	HGB	g/dl	<b>6.2</b>		<b>6.1</b>		± 0.4	<b>13.3</b>		<b>13.5</b>		± 0.5	<b>17.8</b>		<b>18.3</b>		± 0.7												
		g/l	<b>62</b>		<b>61</b>		± 4	<b>133</b>		<b>135</b>		± 5	<b>178</b>		<b>183</b>		± 7												
		mmol/l	<b>3.85</b>		<b>3.79</b>		± 0.25	<b>8.26</b>		<b>8.38</b>		± 0.31	<b>11.05</b>		<b>11.36</b>		± 0.44												
HT	HCT	%	<b>17.5</b>		<b>17.4</b>		± 2.0	<b>36.9</b>		<b>36.8</b>		± 2.5	<b>50.0</b>		<b>49.9</b>		± 3.0												
		l/l	<b>0.175</b>		<b>0.174</b>		± 0.020	<b>0.369</b>		<b>0.368</b>		± 0.025	<b>0.500</b>		<b>0.499</b>		± 0.030												
VGM	MCV	µm <sup>3</sup> ; fl	<b>68</b>		<b>69.0</b>		± 4	<b>80</b>		<b>80.0</b>		± 4	<b>87</b>		<b>87.0</b>		± 4												
TGMH	MCH	pg	<b>24.1</b>		<b>24.2</b>		± 2.4	<b>28.9</b>		<b>29.3</b>		± 2.4	<b>31.0</b>		<b>31.9</b>		± 2.8												
		fmol	<b>1.50</b>		<b>1.50</b>		± 0.15	<b>1.79</b>		<b>1.82</b>		± 0.15	<b>1.92</b>		<b>1.98</b>		± 0.18												
CCMH	MCHC	g/dl	<b>35.5</b>		<b>35.1</b>		± 3.0	<b>36.1</b>		<b>36.7</b>		± 3.0	<b>35.6</b>		<b>36.7</b>		± 3.0												
		g/l	<b>355</b>		<b>351</b>		± 30	<b>361</b>		<b>367</b>		± 30	<b>356</b>		<b>367</b>		± 30												
		mmol/l	<b>22.03</b>		<b>21.79</b>		± 1.86	<b>22.40</b>		<b>22.78</b>		± 1.86	<b>22.10</b>		<b>22.80</b>		± 1.86												
IDR	RDW	%	<b>13.5</b>		<b>12.0</b>		± 3.0	<b>13.0</b>		<b>11.5</b>		± 3.0	<b>12.5</b>		<b>11.0</b>		± 3.0												
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>74</b>		<b>67</b>		± 20	<b>258</b>		<b>252</b>		± 40	<b>510</b>		<b>505</b>		± 60												
VPM	MPV	µm <sup>3</sup> ; fl	<b>8.5</b>		<b>9.3</b>		± 2.0	<b>8.1</b>		<b>8.6</b>		± 2.0	<b>8.1</b>		<b>8.4</b>		± 2.0												
LYMPHO	#		<b>1.3</b>		N/A		± 0.3	<b>2.7</b>		N/A		± 0.5	<b>3.7</b>		N/A		± 1.3												
	%		<b>64.5</b>		N/A		± 8.0	<b>34.0</b>		N/A		± 6.0	<b>18.0</b>		N/A		± 6.0												
MONO	#		<b>0.2</b>		N/A		± 0.2	<b>0.6</b>		N/A		± 0.5	<b>1.1</b>		N/A		± 1.0												
	%		<b>9.0</b>		N/A		± 7.0	<b>7.5</b>		N/A		± 5.0	<b>5.5</b>		N/A		± 5.0												
GRANULO	#		<b>0.5</b>		N/A		± 0.2	<b>4.6</b>		N/A		± 0.6	<b>15.8</b>		N/A		± 1.5												
	%		<b>26.5</b>		N/A		± 7.0	<b>58.5</b>		N/A		± 6.0	<b>76.5</b>		N/A		± 7.0												
CRP	mg/dl		<b>0.54</b>		<b>0.50</b>		± 0.40	<b>2.50</b>		<b>2.50</b>		± 0.80	<b>4.10</b>		<b>3.90</b>		± 1.50												
	mg/l		<b>5.4</b>		<b>5.0</b>		± 4.0	<b>25.0</b>		<b>25.0</b>		± 8.0	<b>41.0</b>		<b>39.0</b>		± 15.0												

Ref: TEMP-1070 Rev.36 BACK/VERSO 9930089-A