

ABX Minotrol CRP



LOT

MC 420

CONTROL



(Exp.) 2021-10-05
(YYYY-MM-DD)

Rev 1

| PARAMETRES PARAMETERS | | UNITES UNITS | | ABX Lysebio | | | | | | | | | | | | | | | | |
|--------------------------|------|---|--------------|--------------|---------|--------------|--------|-------------------------|--------------|--------------|---------|--------------|-------------------------|---------|--------------|--------------|---------|--------------|-------------------------|---------|
| | | | | CONTROL | | | | 1 | CONTROL | | | | 2 | CONTROL | | | | 3 | TOLERANCES TOLERANCE | |
| | | | | MICROS | | PENTRA | | TOLERANCES TOLERANCE | MICROS | | PENTRA | | TOLERANCES TOLERANCE | MICROS | | PENTRA | | | | |
| | | | | EMI CRP | LC-767G | | MS CRP | | | EMI CRP | LC-767G | | | MS CRP | | EMI CRP | LC-767G | | | MS CRP |
| GB | WBC | 10 ³ /mm ³ ;10 ⁹ /l | 2.0 | 2.0 | | 2.1 | | ± 0.4 | 8.0 | 7.9 | | 8.1 | | ± 1.0 | 20.7 | 20.7 | | 21.5 | | ± 2.2 |
| GR | RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 2.40 | 2.39 | | 2.34 | | ± 0.15 | 4.59 | 4.53 | | 4.52 | | ± 0.18 | 5.51 | 5.47 | | 5.52 | | ± 0.25 |
| HB | HGB | g/dl | 6.2 | 6.1 | | 5.9 | | ± 0.4 | 13.6 | 13.5 | | 13.5 | | ± 0.5 | 17.6 | 17.6 | | 17.9 | | ± 0.7 |
| | | g/l | 62 | 61 | | 59 | | ± 4 | 136 | 135 | | 135 | | ± 5 | 176 | 176 | | 179 | | ± 7 |
| | | mmol/l | 3.85 | 3.79 | | 3.66 | | ± 0.25 | 8.45 | 8.38 | | 8.38 | | ± 0.31 | 10.93 | 10.93 | | 11.12 | | ± 0.44 |
| HT | HCT | % | 16.8 | 16.7 | | 16.4 | | ± 2.0 | 37.6 | 37.1 | | 36.8 | | ± 2.5 | 49.0 | 49.2 | | 48.9 | | ± 3.0 |
| | | l/l | 0.168 | 0.167 | | 0.164 | | ± 0.020 | 0.376 | 0.371 | | 0.368 | | ± 0.025 | 0.490 | 0.492 | | 0.489 | | ± 0.030 |
| VGM | MCV | µm ³ ; fl | 70 | 70 | | 70.0 | | ± 4 | 82 | 82 | | 81.5 | | ± 4 | 89 | 90 | | 88.5 | | ± 4 |
| TGMH | MCH | pg | 25.8 | 25.5 | | 25.2 | | ± 2.4 | 29.6 | 29.8 | | 29.9 | | ± 2.4 | 31.9 | 32.2 | | 32.4 | | ± 2.8 |
| | | fmol | 1.60 | 1.58 | | 1.57 | | ± 0.15 | 1.84 | 1.85 | | 1.85 | | ± 0.15 | 1.98 | 2.00 | | 2.01 | | ± 0.18 |
| CCMH | MCHC | g/dl | 36.9 | 36.5 | | 36.0 | | ± 3.0 | 36.1 | 36.3 | | 36.6 | | ± 3.0 | 35.9 | 35.8 | | 36.6 | | ± 3.0 |
| | | g/l | 369 | 365 | | 360 | | ± 30 | 361 | 363 | | 366 | | ± 30 | 359 | 358 | | 366 | | ± 30 |
| | | mmol/l | 22.92 | 22.64 | | 22.37 | | ± 1.86 | 22.44 | 22.57 | | 22.76 | | ± 1.86 | 22.29 | 22.20 | | 22.75 | | ± 1.86 |
| IDR | RDW | % | 14.0 | 13.0 | | 15.5 | | ± 3.0 | 12.5 | 12.5 | | 13.5 | | ± 3.0 | 12.0 | 12.0 | | 12.5 | | ± 3.0 |
| PLAQ. | PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 73 | 79 | | 66 | | ± 20 | 262 | 260 | | 248 | | ± 40 | 490 | 490 | | 495 | | ± 60 |
| VPM | MPV | µm ³ ; fl | 8.5 | 9.1 | | 10.1 | | ± 2.0 | 8.8 | 8.7 | | 9.3 | | ± 2.0 | 8.5 | 8.5 | | 9.0 | | ± 2.0 |
| LYMPHO | | # | 1.3 | 1.3 | | N/A | | ± 0.2 | 3.0 | 2.8 | | N/A | | ± 0.5 | 3.5 | 3.3 | | N/A | | ± 1.1 |
| | | % | 67.0 | 66.0 | | N/A | | ± 8.0 | 37.0 | 36.0 | | N/A | | ± 6.0 | 17.0 | 16.0 | | N/A | | ± 5.0 |
| MONO | | # | 0.2 | 0.2 | | N/A | | ± 0.2 | 0.4 | 0.5 | | N/A | | ± 0.4 | 1.3 | 1.3 | | N/A | | ± 1.3 |
| | | % | 5.5 | 6.5 | | N/A | | ± 5.5 | 5.5 | 6.0 | | N/A | | ± 5.5 | 6.5 | 6.5 | | N/A | | ± 6.5 |
| GRANULO | | # | 0.6 | 0.6 | | N/A | | ± 0.2 | 4.6 | 4.6 | | N/A | | ± 0.5 | 15.8 | 16.0 | | N/A | | ± 1.5 |
| | | % | 27.5 | 27.5 | | N/A | | ± 7.0 | 57.5 | 58.0 | | N/A | | ± 6.0 | 76.5 | 77.5 | | N/A | | ± 7.0 |
| CRP | | mg/dl | 0.73 | 0.83 | | 0.63 | | ± 0.40 | 2.48 | 2.90 | | 2.21 | | ± 0.80 | 4.29 | 4.88 | | 4.03 | | ± 1.50 |
| | | mg/l | 7.3 | 8.3 | | 6.3 | | ± 4.0 | 24.8 | 29.0 | | 22.1 | | ± 8.0 | 42.9 | 48.8 | | 40.3 | | ± 15.0 |

Ref: TEMP-1070 Rev.39 BACKVERSO 1300092170