

ABX Minotrol Retic



LOT RX 419
Rev 1

CONTROL

(Exp.) 2019-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 0.8	4.8	4.8	4.8	± 2.1	10.5	10.5	10.5	± 3.2
RETIC	%	1.44	1.64	1.64	± 0.82	4.72	5.20	5.20	± 1.30	10.00	10.90	10.90	± 2.20
RETIC	10 ⁶ /µl	0.049	0.056	0.056	± 0.028	0.160	0.176	0.176	± 0.044	0.343	0.375	0.375	± 0.075
MFI MRI *		30.0	27.9	27.9	± 8.0	43.0	41.3	41.3	± 10.0	46.2	44.0	44.0	± 12.0
GR RBC	10 ⁶ /µl	3.42	3.42	3.42	± 0.30	3.38	3.39	3.39	± 0.30	3.43	3.44	3.44	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 419
Rev 1

CONTROL

(Exp.) 2019-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.5			± 0.8	4.8			± 2.1	10.5			± 3.2
RETIC	%	1.62			± 0.81	5.14			± 1.29	10.65			± 2.15
RETIC	10 ⁶ /µl	0.055			± 0.028	0.172			± 0.043	0.362			± 0.072
MFI MRI *		24.9			± 8.0	39.6			± 10.0	43.8			± 12.0
GR RBC	10 ⁶ /µl	3.40			± 0.30	3.35			± 0.30	3.40			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA