

ABX Minotrol Retic



LOT RX 410
Rev 1

CONTROL

(Exp.) 2018-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 0.8	6.1	6.1	6.1	± 2.4	10.5	10.5	10.5	± 3.2
RETIC	%	1.80	1.94	1.94	± 0.97	5.59	5.91	5.91	± 1.48	10.70	11.20	11.20	± 2.30
RETIC	10 ⁶ /µl	0.062	0.067	0.067	± 0.034	0.195	0.206	0.206	± 0.052	0.351	0.373	0.373	± 0.075
MFI MRI *		34.7	31.8	31.8	± 8.0	51.7	51.1	51.1	± 10.0	55.5	55.6	55.6	± 12.0
GR RBC	10 ⁶ /µl	3.45	3.46	3.46	± 0.30	3.48	3.48	3.48	± 0.30	3.28	3.33	3.33	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 410
Rev 1

CONTROL

(Exp.) 2018-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA		1		PENTRA		2		PENTRA		3	
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.5			± 0.8	6.1			± 2.4	10.5			± 3.2
RETIC	%	1.87			± 0.94	5.78			± 1.45	11.10			± 2.22
RETIC	10 ⁶ /µl	0.065			± 0.033	0.202			± 0.051	0.370			± 0.074
MFI MRI *		31.2			± 8.0	49.7			± 10.0	53.5			± 12.0
GR RBC	10 ⁶ /µl	3.47			± 0.30	3.49			± 0.30	3.33			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA