

# ABX Minotrol Retic



**LOT** RX 420  
Rev 1

**CONTROL**

(Exp.) 2020-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.4	1.4	1.4	± 0.7	5.3	5.3	5.3	± 2.1	11.3	11.3	11.3	± 3.4
RETIC	%	1.80	2.00	2.00	± 1.00	5.19	5.59	5.59	± 1.40	10.40	11.10	11.10	± 2.30
RETIC	10 <sup>6</sup> /µl	0.062	0.069	0.069	± 0.035	0.180	0.195	0.195	± 0.049	0.357	0.382	0.382	± 0.076
MFI MRI *		34.1	34.4	34.4	± 8.0	43.9	44.6	44.6	± 10.0	47.2	48.8	48.8	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.44	3.45	3.45	± 0.30	3.46	3.48	3.48	± 0.30	3.43	3.44	3.44	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 420  
Rev 1

**CONTROL**

(Exp.) 2020-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.4</b>			± 0.7	<b>5.3</b>			± 2.1	<b>11.3</b>			± 3.4
RETIC	%	<b>1.96</b>			± 0.98	<b>5.49</b>			± 1.37	<b>10.90</b>			± 2.18
RETIC	10 <sup>6</sup> /µl	<b>0.068</b>			± 0.034	<b>0.191</b>			± 0.048	<b>0.374</b>			± 0.075
MFI MRI *		<b>35.4</b>			± 8.0	<b>45.7</b>			± 10.0	<b>48.7</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.45</b>			± 0.30	<b>3.48</b>			± 0.30	<b>3.43</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

\* For research use only in Canada and USA