

ABX Minotrol Retic



LOT RX 401
Rev 1

CONTROL

(Exp.) 2016-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.8	1.8	1.8	± 1.0	6.1	6.1	6.1	± 2.4	11.0	11.0	11.0	± 3.3
RETIC	%	2.14	2.15	2.15	± 1.08	5.69	5.78	5.78	± 1.45	11.50	11.50	11.50	± 2.30
RETIC	10 ⁶ /µl	0.075	0.075	0.075	± 0.038	0.197	0.200	0.200	± 0.050	0.397	0.394	0.394	± 0.079
MFI MRI *		27.6	27.3	27.3	± 8.0	47.7	46.7	46.7	± 10.0	49.9	48.9	48.9	± 12.0
GR RBC	10 ⁶ /µl	3.51	3.50	3.50	± 0.30	3.47	3.46	3.46	± 0.30	3.45	3.43	3.43	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 401
Rev 1

CONTROL

(Exp.) 2016-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1		2		3							
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
XLR			XLR			XLR							
Méthode Manuelle / Manual RETIC	%	1.8			± 1.0	6.1			± 2.4	11.0			± 3.3
RETIC	%	2.14			± 1.07	5.61			± 1.40	11.30			± 2.26
RETIC	10 ⁶ /µl	0.075			± 0.038	0.195			± 0.049	0.390			± 0.078
MFI MRI *		26.7			± 8.0	45.6			± 10.0	47.6			± 12.0
GR RBC	10 ⁶ /µl	3.52			± 0.30	3.47			± 0.30	3.45			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A

* For research use only in Canada and USA