

ABX Minotrol Retic



LOT RX 412
Rev 1

CONTROL

(Exp.) 2018-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.7	1.7	1.7	± 0.9	5.5	5.5	5.5	± 2.2	10.2	10.2	10.2	± 3.1
RETIC	%	2.20	2.21	2.21	± 1.11	5.25	5.49	5.49	± 1.37	10.60	11.10	11.10	± 2.30
RETIC	10 ⁶ /µl	0.077	0.076	0.076	± 0.039	0.185	0.189	0.189	± 0.047	0.373	0.384	0.384	± 0.077
MFI MRI *		32.0	32.2	32.2	± 8.0	43.3	43.3	43.3	± 10.0	47.6	48.9	48.9	± 12.0
GR RBC	10 ⁶ /µl	3.48	3.43	3.43	± 0.30	3.52	3.45	3.45	± 0.30	3.52	3.46	3.46	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 412
Rev 1

CONTROL

(Exp.) 2018-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA		1		PENTRA		2		PENTRA		3	
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.7			± 0.9	5.5			± 2.2	10.2			± 3.1
RETIC	%	2.07			± 1.04	5.35			± 1.34	11.00			± 2.20
RETIC	10 ⁶ /µl	0.070			± 0.035	0.183			± 0.046	0.375			± 0.075
MFI MRI *		32.5			± 8.0	41.8			± 10.0	46.0			± 12.0
GR RBC	10 ⁶ /µl	3.39			± 0.30	3.42			± 0.30	3.41			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA