

ABX Minotrol Retic



LOT RX 036
Rev 1

CONTROL

(Exp.) 2016-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.1	1.1	1.1	± 0.6	5.5	5.5	5.5	± 2.2	11.0	11.0	11.0	± 3.3
RETIC	%	1.98	2.01	2.01	± 1.01	5.41	5.65	5.65	± 1.41	11.20	11.30	11.30	± 2.30
RETIC	10 ⁶ /µl	0.068	0.069	0.069	± 0.035	0.186	0.192	0.192	± 0.048	0.385	0.383	0.383	± 0.077
MFI MRI *		28.4	27.0	27.0	± 8.0	44.7	44.3	44.3	± 10.0	47.6	47.6	47.6	± 12.0
GR RBC	10 ⁶ /µl	3.44	3.41	3.41	± 0.30	3.43	3.39	3.39	± 0.30	3.44	3.39	3.39	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 036
Rev 1

CONTROL

(Exp.) 2016-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.1			± 0.6	5.5			± 2.2	11.0			± 3.3
RETIC	%	2.04			± 1.02	5.62			± 1.41	11.60			± 2.40
RETIC	10 ⁶ /µl	0.071			± 0.036	0.195			± 0.049	0.399			± 0.080
MFI MRI *		26.7			± 8.0	43.4			± 10.0	46.2			± 12.0
GR RBC	10 ⁶ /µl	3.47			± 0.30	3.47			± 0.30	3.44			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A

* For research use only in Canada and USA