

ABX Minotrol Retic



LOT RX 404
Rev 2

CONTROL

(Exp.) 2017-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.7	1.7	1.7	± 0.9	5.5	5.5	5.5	± 2.2	12.0	12.0	12.0	± 4.2
RETIC	%	1.95	2.13	2.13	± 1.07	5.38	5.48	5.48	± 1.37	11.00	11.00	11.00	± 2.20
RETIC	10 ⁶ /µl	0.067	0.073	0.073	± 0.037	0.182	0.184	0.184	± 0.046	0.372	0.367	0.367	± 0.074
MFI MRI *		27.2	27.6	27.6	± 8.0	43.4	44.9	44.9	± 10.0	46.3	49.4	49.4	± 12.0
GR RBC	10 ⁶ /µl	3.45	3.44	3.44	± 0.30	3.39	3.36	3.36	± 0.30	3.38	3.34	3.34	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 404
Rev 2

CONTROL

(Exp.) 2017-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.7			± 0.9	5.5			± 2.2	12.0			± 4.2
RETIC	%	1.89			± 0.95	5.28			± 1.32	11.00			± 2.20
RETIC	10 ⁶ /µl	0.067			± 0.034	0.181			± 0.045	0.375			± 0.075
MFI MRI *		28.8			± 8.0	45.3			± 10.0	48.7			± 12.0
GR RBC	10 ⁶ /µl	3.52			± 0.30	3.42			± 0.30	3.41			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA