

# ABX Minotrol Retic



**LOT** RX 405  
Rev 1

**CONTROL**

(Exp.) 2017-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 0.8	6.0	6.0	6.0	± 2.4	11.7	11.7	11.7	± 3.5
RETIC	%	1.97	2.10	2.10	± 1.05	5.62	5.66	5.66	± 1.42	12.00	11.90	11.90	± 2.40
RETIC	10 <sup>6</sup> /µl	0.068	0.072	0.072	± 0.036	0.194	0.194	0.194	± 0.049	0.410	0.403	0.403	± 0.082
MFI MRI *		25.9	27.7	27.7	± 8.0	43.5	45.7	45.7	± 10.0	47.9	49.7	49.7	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.44	3.43	3.43	± 0.30	3.46	3.42	3.42	± 0.30	3.42	3.39	3.39	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 405  
Rev 1

**CONTROL**

(Exp.) 2017-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.5</b>			± 0.8	<b>6.0</b>			± 2.4	<b>11.7</b>			± 3.5
RETIC	%	<b>2.01</b>			± 1.01	<b>5.55</b>			± 1.39	<b>11.80</b>			± 2.40
RETIC	10 <sup>6</sup> /µl	<b>0.070</b>			± 0.035	<b>0.192</b>			± 0.048	<b>0.405</b>			± 0.081
MFI MRI *		<b>27.3</b>			± 8.0	<b>45.4</b>			± 10.0	<b>50.4</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.50</b>			± 0.30	<b>3.46</b>			± 0.30	<b>3.43</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

\* For research use only in Canada and USA