

ABX Minotrol Retic



LOT RX 425
Rev 1

CONTROL

(Exp.) 2020-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.3	1.3	1.3	± 1.0	5.1	5.1	5.1	± 2.1	10.9	10.9	10.9	± 3.3
RETIC	%	1.30	1.49	1.49	± 0.75	4.53	4.90	4.90	± 1.23	9.90	10.70	10.70	± 2.20
RETIC	10 ⁶ /µl	0.045	0.052	0.052	± 0.026	0.150	0.166	0.166	± 0.042	0.340	0.370	0.370	± 0.074
MFI MRI *	%	27.3	26.0	26.0	± 8.0	40.2	41.0	41.0	± 10.0	46.6	47.5	47.5	± 12.0
GR RBC	10 ⁶ /µl	3.44	3.48	3.48	± 0.30	3.32	3.38	3.38	± 0.30	3.43	3.46	3.46	± 0.30
PIC RANGE	-	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.38 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 425
Rev 1

CONTROL

(Exp.) 2020-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.3			± 1.0	5.1			± 2.1	10.9			± 3.3
RETIC	%	1.63			± 0.82	4.97			± 1.24	10.60			± 2.20
RETIC	10 ⁶ /µl	0.055			± 0.028	0.163			± 0.041	0.358			± 0.072
MFI MRI *	%	22.9			± 8.0	41.0			± 10.0	46.8			± 12.0
GR RBC	10 ⁶ /µl	3.40			± 0.30	3.28			± 0.30	3.38			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.38 BACK / VERSO 1219930083-A

* For research use only in Canada and USA