

# ABX Minotrol Retic



**LOT** RX 428  
Rev 1

**CONTROL**

(Exp.) 2021-05-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
		DX120	DX Nexus		DX120	DX Nexus		DX120	DX Nexus		DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%		1.6	1.6	± 1.0		5.6	5.6	± 2.2		11.3	11.3	± 3.4
RETIC	%		1.58	1.58	± 0.79		5.00	5.00	± 1.25		10.60	10.60	± 2.10
RETIC	10 <sup>6</sup> /µl		0.053	0.053	± 0.027		0.169	0.169	± 0.042		0.360	0.360	± 0.072
MFI MRI *	%		29.7	29.7	± 8.0		46.0	46.0	± 10.0		50.3	50.3	± 12.0
GR RBC	10 <sup>6</sup> /µl		3.35	3.35	± 0.30		3.38	3.38	± 0.30		3.40	3.40	± 0.30
PIC RANGE	-		27.0	27.0	± 5.0		27.0	27.0	± 5.0		27.0	27.0	± 5.0

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.38 FRONT / RECTO 1219930083-A

# ABX Minotrol Retic



**LOT** RX 428  
Rev 1

**CONTROL**

(Exp.) 2021-05-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.6</b>			± 1.0	<b>5.6</b>			± 2.2	<b>11.3</b>			± 3.4
RETIC	%	<b>1.71</b>			± 0.86	<b>5.10</b>			± 1.28	<b>10.70</b>			± 2.10
RETIC	10 <sup>6</sup> /µl	<b>0.057</b>			± 0.029	<b>0.170</b>			± 0.043	<b>0.358</b>			± 0.072
MFI MRI *	%	<b>25.4</b>			± 8.0	<b>42.3</b>			± 10.0	<b>48.6</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.31</b>			± 0.30	<b>3.33</b>			± 0.30	<b>3.35</b>			± 0.30
PIC RANGE	-	<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.38 BACK / VERSO 1219930083-A

\* For research use only in Canada and USA