

ABX Minotrol Retic



LOT RX 056
Rev 1

CONTROL

(Exp.) 2016-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.6	1.6	1.6	± 0.8	5.5	5.5	5.5	± 2.2	10.2	10.2	10.2	± 3.1
RETIC	%	2.10	2.11	2.11	± 1.06	5.66	5.66	5.66	± 1.42	11.70	11.60	11.60	± 2.40
RETIC	10 ⁶ /µl	0.073	0.072	0.072	± 0.037	0.195	0.195	0.195	± 0.049	0.401	0.394	0.394	± 0.082
MFI MRI *		30.6	31.0	31.0	± 8.0	44.7	44.8	44.8	± 10.0	49.2	50.6	50.6	± 12.0
GR RBC	10 ⁶ /µl	3.47	3.43	3.43	± 0.30	3.45	3.45	3.45	± 0.30	3.43	3.40	3.40	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 056
Rev 1

CONTROL

(Exp.) 2016-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.6			± 0.8	5.5			± 2.2	10.2			± 3.1
RETIC	%	2.00			± 1.00	5.47			± 1.37	11.30			± 2.26
RETIC	10 ⁶ /µl	0.069			± 0.035	0.187			± 0.047	0.385			± 0.077
MFI MRI *		28.9			± 8.0	43.1			± 10.0	49.1			± 12.0
GR RBC	10 ⁶ /µl	3.43			± 0.30	3.41			± 0.30	3.41			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A

* For research use only in Canada and USA